Meridian Health Jersey Shore University Medical Center Core Laboratory Neptune, NJ 07753

SUBJECT: QUALITY MANAGEMENT		TITLE: COMPARABILITY OF CLINITEK URINALYSIS ANALYZERS	
EFFECTIVE DATE: 12/20/2015		SECTION: URINALYSIS SOP	
APPROVED:		POSITION RESPONSIBLE FOR REVIEW: Laboratory Director or Designee	
PURPOSE:	To provide a process for the documentation of comparability between the Clinitek Atlas and the Clinitek Status urine analyzers. The analyzers are checked against each other at least twice a year for comparability of results.		
RESPONSIBILITY:	Laboratory Supervisors, Technical Staff		
PROCEDURE:	a folder in which 1. 2. 3. 4.	For each test note the number of samples that were exact match, those that were Within one grade of each other, and those that exceeded one grade. Example: TEST Exact match One grade two or more grades Blood 9/10 10/10 0/10 Nitrite 10/10 10/10 0/10 Log results on the Urinalysis Comparability form Any results 2 or more grades apart must be investigated.	

Created On: 12/20/2015 By: Richard Hull MT(ASCP)	Revised:	Reviewed:
Reviewed:	Reviewed:	Reviewed:

Meridian Health Jersey Shore University Medical Center Core Laboratory Neptune, NJ 07753