

MEDTOX Scan Urine Drugs of Abuse Panel Manual Log Meter #1 SN 1018M1440

Reprint Acc label after performing and verifying results. Reprinted label should have reflexed tests.




Benzo and Barb will not be on your reflex label. It will need to be acc in KRMS for send out to Quest.

NOTE: If interface is down 2nd tech verif. required prior to reprinting.

DATE	PATIENT NAME	Date sent out	CTRL1	THC	PCP	COC	CTRL2	mAMP	OPI	CTRL3	AMP	BZO	TCA	CTRL4	BAR	TECH/ 2nd TECH

Indicate in CTRL column if test is VALID. If test is INVALID DO NOT REPORT, repeat test.

Document History Page

Change type: New, Major, Minor etc.	Changes Made to Document – describe	Signature responsible person/date	Lab Operations Director Review/Date	Laboratory Medical Director Review/ Date	Date change implemented
Major	<ul style="list-style-type: none"> LCS2108-A assigned to Form MEDTOXScan Urine Drugs of Abuse Panel Manual Log for Meter#1 Name of form and from number designation appears in the footer section of the form Process or procedure associated with for is Profile®-V MEDTOXScan® Sample Porcessing for Drugs of Abuse Test System LCS 2108. 	R.Castaneto 08-23-16	J.Wolf 08-26-16	S.Wirio M.D. 08-26-16	08-26-16
Minor	<ul style="list-style-type: none"> Form updated to reflect current Meter #1 Instrument Serial Number. Form LCS 2108-#1B. 	 7/30/18	 7-31-18	 8/5/18	8/5/18

