

URINALYSIS QC ACTION LOG

Core Lab

Time/Date: 22:00 / 8/11/17 Control CA Tech FC

Problem/Outlier: Bilirubin - Failed

Corrective Action: Rerun using same aliquot of CA control -
QC Passed.

Patient Results Reported: Yes / No
Supervisor Signature: _____ Date: _____

Time/Date: 21:15 8/12/17 Control CB Tech FC

Problem/Outlier: Blood - Failed

Corrective Action: Rerun QC using same aliquot of CB control -
still failed. Repeat run using fresh aliquot of
QC material - QC Passed.

Patient Results Reported: Yes / No
Supervisor Signature: _____ Date: _____

Time/Date: 20:00 8/15/17 Control IQ Positive QC Tech FC

Problem/Outlier: IQ Pos. Control - Failed

Corrective Action: Rerun using same aliquot - still failed.
Repeat run using fresh aliquot of IQ Pos. QC - still failed.
Repeat run using new bottle of IQ Pos. QC - QC Passed.

Patient Results Reported: Yes / No
Supervisor Signature: _____ Date: _____