

Beaumont

Origination 3/3/2020
Last Approved 4/11/2024
Effective 4/11/2024
Last Revised 4/11/2024
Next Review 4/11/2026

Document Contact Christopher Ferguson: Dir, Lab Operations B
Area Laboratory-Safety
Applicability All Beaumont Hospitals

Dry Ice Handling

Document Type: Procedure

I. PURPOSE AND OBJECTIVE:

The purpose of this procedure is to provide Corewell Health Laboratory employees a guide to safely use, store, package, transport and dispose of dry ice. The objective is to reduce risk to employees handling dry ice. Dry ice is considered a hazard for three reasons:

- A. **Explosion hazard:** Dry ice releases a large volume of carbon dioxide gas as it sublimates (solid vaporizes to gas state). If packaged in a container that does not allow for the release of the gas, it may explode, causing personal injury or property damage.
- B. **Suffocation hazard:** a large volume of carbon dioxide gas emitted in a confined space (e.g. transport vehicle with recirculated air), or other unventilated area (e.g. cold room) may create an oxygen deficient atmosphere.
- C. **Contact hazard:** Dry ice is a cryogenic material that causes severe frostbite upon contact with the skin or eyes.

II. PROCEDURE:

- A. Safe use of dry ice
 - 1. Avoid eye or skin contact.
 - a. NEVER handle dry ice with bare hands.
 - b. When handling dry ice use insulated gloves and dry ice tongs or plastic scoop.
 - c. Always use appropriate eye protection (e.g. goggles or face shield).
 - 2. Avoid use of dry ice in confined areas such as walk-in refrigerators, environmental

chambers or rooms without ventilation. A leak in such areas could cause an oxygen-deficient atmosphere.

B. Safe storage of dry ice

1. Do not store dry ice in confined areas such as walk-in refrigerators, environmental chambers or rooms without ventilation. A leak in such areas could cause oxygen-deficient atmosphere.
2. Never place dry ice on tile or laminated surfaces because adhesive will be destroyed.
3. Never store dry ice in a sealed, airtight container at a temperature above the boiling point of the dry ice (-109°F); the pressure resulting from production of carbon dioxide may lead to explosion.

C. Safe packaging of dry ice

1. Dry ice is placed in an insulated container such as Styrofoam inserts and not placed directly in a cardboard box.
2. Samples should be surrounded by an absorbent material and then placed in a secondary containment.
3. Secure samples in a way when dry ice sublimates, samples will not move freely inside insulated box.
4. Minimize the volume of air dry ice is exposed in order to slow rate of sublimation. If there is any air space after you fill your package with dry ice, fill it with packaging peanuts or other material to reduce the volume of air space.
5. Outer packing (i.e. cardboard box) should be free from damage. Do not use plastics or other brittle material for outer package.
6. Tape box in a way that will not interfere with gas release (e.g. Do not fully tape seams of the box).

D. Safe transport of dry ice

1. Packages should be transported in a secure position in a well-ventilated vehicle.
2. Product transported in an enclosed, non-ventilated compartment of a vehicle can present serious safety hazards. (e.g. Do not have vehicle ventilation control set to recirculate the air). If it is transported inside a car or van for more than 15 minutes make sure there is fresh air.

E. Safe disposal of dry ice

1. Dispose unneeded dry ice by letting the unused portion sublimate in a well-ventilated area.
2. NEVER dispose dry ice in a sink, toilet, or other device. Such action can destroy the structure because of the temperature difference.
3. NEVER dispose of dry ice in the trash or garbage.
4. NEVER leave unneeded dry ice in hallways or other public areas.

III. FIRST AID:

A. Skin Contact

1. Prolonged dry ice contact with the skin will freeze cells and cause injury similar to a burn. In case of exposure to dry ice, remove any clothing that is not frozen to the skin.
2. Do not rub frozen body parts because tissue damage may result.
3. Place the affected body part in a warm bath (not above 104°F). NEVER use dry heat.
4. Obtain medical assistance as soon as possible.
5. Submit a SafetyPause through The Well intranet (click on *Report a Safety Concern* on The Well homepage, followed by *Submit a SafetyPause*).

B. Eye Contact

1. Immediately and briefly flush with lukewarm gently flowing water. DO NOT attempt to rewarm. Cover both eyes with sterile dressing. Treatment is urgently required.
2. Submit a SafetyPause through The Well intranet (click on *Report a Safety Concern* on The Well homepage, followed by *Submit a SafetyPause*).

C. Inhalation

1. Take precautions to secure your own safety before attempting to rescue. Treatment is urgently required.
2. Submit a SafetyPause through The Well intranet (click on *Report a Safety Concern* on The Well homepage, followed by *Submit a SafetyPause*).

IV. REFERENCES:

- A. Occupational Safety and Health Administration (OSHA) Laboratory Safety Guidance, www.osha.gov, OSHA 3404-11R 2011
- B. [Safety Data Sheets - SDS Vault](#)

Approval Signatures

Step Description	Approver	Date
CLIA Site Licensed Medical Directors	Muhammad Arshad: Chief, Pathology	4/11/2024
CLIA Site Licensed Medical Directors	Ann Marie Blenc: System Med Dir, Hematopath	4/11/2024
CLIA Site Licensed Medical Directors	Masood Siddiqui: Staff Pathologist	4/11/2024

CLIA Site Licensed Medical Directors	Hassan Kanaan: OUWB Clinical Faculty	4/10/2024
CLIA Site Licensed Medical Directors	Subhashree Mallika Krishnan: Staff Physician	4/10/2024
CLIA Site Licensed Medical Directors	Kurt Bernacki: System Med Dir, Surgical Path	4/10/2024
CLIA Site Licensed Medical Directors	Ryan Johnson: OUWB Clinical Faculty	4/10/2024
CLIA Site Licensed Medical Directors	Jeremy Powers: Chief, Pathology	4/10/2024
CLIA Site Licensed Medical Directors	John Pui: Chief, Pathology	4/10/2024
Policy and Forms Steering Committee Approval (if needed)	Christopher Ferguson: Dir, Lab Operations B	4/9/2024
	Sarah Britton: VP, Laboratory Svcs	4/9/2024
Operations Directors	Joan Wehby: Dir, Lab Operations C	4/9/2024
Operations Directors	Brittnie Berger: Dir Sr, Lab Operations	4/2/2024
Operations Directors	Elzbieta Wystepek: Dir, Lab Operations B	3/11/2024
Operations Directors	Amy Knaus: Dir, Lab Operations C	3/9/2024
Operations Directors	Christopher Ferguson: Mgr, Laboratory	3/8/2024
	Christopher Ferguson: Mgr, Laboratory	3/8/2024

Applicability

Dearborn, Farmington Hills, Grosse Pointe, Royal Oak, Taylor, Trenton, Troy, Wayne