

**Central Processing RL6s August 2022**

8/1/2022	UDS received in UA tube. Not poured off, put on UA analyzer, unable to use for UDS testing.
8/1/2022	Sendout test canceled in CP. Not reordered. Had to be redrawn.
8/7/2022	Spun CBC
8/7/2022	Spun CBC
8/8/2022	Wrong swab types used for Covid PCR test kits. Multiple tests canceled on same patient.
8/9/2022	Ammonia sample dumped in Spin trash; found past stability
8/13/2022	Bags of samples found in accessioning bin on top shelf 2 days after collection. All samples expired and required recollection. 27 patients impacted.
8/14/2022	Spun CBC
8/15/2022	Spun CBC
8/15/2022	ED samples not received in lab
8/18/2022	CBC lost between CP and heme
8/19/2022	Rituxan screen spun
8/24/2022	Flow sample spun
8/24/2022	Unspun sample placed in spun rack and run/resulted. Results had to be corrected after sample pulled and centrifuged.