

Central Processing RL6s - October 2022

10/3/2022	CBC spun - gold and lav labels on wrong tubes, not caught in CP
10/3/2022	CBC and CMP labels switched in CP
10/4/2022	CBC spun - gold and lav labels on wrong tubes, not caught in CP
10/17/2022	Unspun sample in spun rack
10/18/2022	Short sample on the chemistry track
10/18/2022	Short sample on the chemistry track
10/18/2022	Short sample on the chemistry track
10/18/2022	Unspun sample in spun rack
10/20/2022	JAK2 sample centrifuged
10/25/2022	Peds ED tests not processed when sample with chart label received
10/26/2022	ED UA sample lost between CP and Hem
10/29/2022	UPC add-on from DMC Lab - correct collection date/time not entered
10/31/2022	CBC spun while track down
10/31/2022	CBC spun while track down