

## **Sweat Chloride Assay Record**

	Test: Initial Repeat	
Patient Identification Label	Limb Used	
Vial Identification		
POST Collection Weightg		
PRE Collection Weightg		
Weight of Sweat Collectedg		
CHLORIDOMETER READINGS		
Low Standard (80 mEq/L) (Range: 76 - 84)		mEq/L
High Standard (120 mEq/L) (Range: 115 - 1	25)	mEq/L
Result for Serum Controls		
Level 1 Control		mEq/L
Level 3 Control		mEq/L
Chloride Concentration of Eluate		mEq/L
Chloridometer Reading (X4 dilution) to verify Gibson Cooke Chloride value greater than 40 mEq/L		mEq/L X4 =mEq/l
QUANTITATIVE SWEAT CHLORID	E VALUE (GIBSON C	OOKE)
Eluate mEq/L g of Sweat Co	X 0.2 = Sweat Chloride rollected	nEq/L
Quantitative Sweat Chloride Value (Gibson <b>NOTE:</b> Do not report values >160 mEq/L. C		mEq/L Patient must be retested.
Technologist	The results in this batch w	ere Second Clerical Check
Date	checked for clerical errors.  By:	Ву:
	Date: Time:	Date: Time: