

# Beaumont

Origination: 1/12/2022  
Effective: 2/17/2022  
Last Approved: 2/17/2022  
Last Revised: 2/17/2022  
Next Review: 2/17/2024

Document Contact: [Colette Kessler: Mgr](#)  
[Laboratory](#)

Area: [Laboratory-Chemistry](#)

Key Words:

Applicability: [Royal Oak](#)

## Automated Chemistry Criteria for Unacceptable Specimens - Royal Oak

Document Type: Procedure

### I. PURPOSE AND OBJECTIVE:

The following list constitutes reasons for which a specimen may be considered unacceptable.

### II. PROCEDURE:

- A. If it is decided that a specimen is unacceptable, the floor or the physician should be notified, and arrangements made to draw another specimen.
- B. If the physician insists that the specimen be tested, perform the test, provided the volume is adequate to do so, and note on the report as to why the result may be unreliable.

### III. REASONS FOR UNACCEPTABLE SPECIMENS:

- A. If an improperly or unlabeled specimen is received, see Clinical Pathology Procedure, [Correction of Information on Specimen Labels: Proper Handling of Unlabeled/Mislabeled Specimens](#)
- B. **Problems related to patient preparation or condition.**

Problem	Test Affected	Corrective Action
Interfering Medication Parenteral Iron (e.g. Iron Sucrose)	Iron	See Comment in report field.
Odd color serum	Variable	Consult with pathologist/clinical chemist about possible interference.

- C. **Problems related to sample collection, preservation or treatment.**

Problem	Test Affected	Corrective Action
Inadequate volume	Any	Redraw

Large air bubbles	Blood Gases	Notify floor to redraw
Not iced specimen	Blood Gases Ammonia	Request redraw
Improper collection tube-serum	Ammonia	Draw in EDTA tube
Oxalate, EDTA	Ca, Mg, Enzymes Others possible	Redraw
Not acidified	Urinary Catecholamines VMA Metanephrines	Acidify to pH 1-3; if initial pH does not exceed 7.0
Improper times or inadequate collection	All clearance tests Urine amylase Other urine tests requiring 24 hour collection	Obtain actual collection time or recollect

D. **Problems related to the sample itself.**

Problem	Test Affected	Corrective Action
Hemolysis >50 mg/dL	PLACA, Ammonia	Redraw
Hemolysis >100 mg/dL	Iron	Redraw
Hemolysis >200 mg/dL	Acetaminophen Lactic Acid Vitamin B12 Folate	Redraw
Hemolysis (in vitro) >500 mg/dL	All tests <b>except:</b> <b>Glucose</b> <b>Creatinine</b> <b>ETOH, BUN</b>	Redraw* Creatinine, ETOH, Glucose, BUN: report with CHVT comment
Hemolysis >1000 mg/dL	Glucose	Redraw*
Gross Hemolysis	Creatinine, ETOH	Report with CHVT comment
Gross Lipemia	Variable, but affects most tests	Ultracentrifuge specimen for all tests except: Cholesterol, HDL Cholesterol, Direct LDL Cholesterol and Triglycerides. Perform Cholesterol, HDL Cholesterol, Direct LDL Cholesterol and Triglycerides on unspun serum.
<b>*If it is determined that in-vivo hemolysis is occurring, testing can be performed. Refer to <a href="#">“Policy for the Determination of Hemolysis, Lipemia and Icterus”</a></b>		

## Attachments

No Attachments

## Approval Signatures

Step Description	Approver	Date
Medical Director	Ann Marie Blenc: System Med Dir, Hematopath	2/17/2022
Policy and Forms Steering Committee Approval (if needed)	Colette Kessler: Mgr Laboratory [RC]	2/9/2022
Policy and Forms Steering Committee Approval (if needed)	Gail Juleff: Project Mgr Policy	2/9/2022
Lab Chemistry Best Practice Committee	Qian Sun: Tech Dir, Clin Chemistry, Path	2/8/2022
Lab Chemistry Best Practice Committee	Elizabeth Sykes: System Med Dir, Chemistry	2/1/2022
	Colette Kessler: Mgr Laboratory [RC]	2/1/2022

## Applicability

Royal Oak