

Beaumont

Origination 12/27/2022
Last Approved 2/6/2024
Effective 2/6/2024
Last Revised 12/27/2022
Next Review 2/5/2026

Document Contact Kelly Walewski:
Supv, Laboratory
Area Laboratory-
Chemistry
Applicability All Beaumont
Hospitals

Laboratory Chemistry Total Error Allowable

Document Type: Procedure

I. PURPOSE AND OBJECTIVE:

- A. This document identifies Total Error Allowable (TEA) is the amount of error that can be tolerated without invalidating the medical usefulness of the analytical result. U.S.Clinical Laboratory Improvement Amendment (CLIA) requirements for a Quality Control (QC) program call for in-house data to conduct a risk assessment of the testing process.

II. BACKGROUND:

- A. Clinical Laboratory Improvement Act (CLIA) requirements for a Quality Control (QC) program call for in-house data to conduct a risk assessment of the testing process. To conduct a risk assessment, the laboratory must identify the sources of potential failures and errors for a testing process, and evaluate the frequency and impact of those failures and sources of error.
- B. In-house data, established by the laboratory in its own environment and by its own personnel, must be included to demonstrate that the stability of the test system supports the number and frequency of QC documented in the Quality Control Plan (QCP). Data from verification or establishment of performance specifications and historical QC data can be included. Published data or data from manufacturers (e.g., package inserts) may be taken into consideration, but may not be used as the sole criteria for decision-making.
- C. The concept of "total error allowable" was investigated by Westgard over a quarter century ago for use in laboratory medicine; it comprises bias as well as random elements.
- D. Bias: Inaccuracy or systematic error is determined during method validation studies from a comparison of methods experiment. Laboratories perform these experiments to verify a manufacturer's claim after installation of new test systems. After initial validation, labs are required to monitor bias using EQA (external quality assessment)/Patient samples with target values established by reference methods, the mean of a survey group or the mean of a survey

peer group.

- E. Imprecision (random error) is determined from a replication experiment during method validation studies or SQC (statistical quality control) data collected during routine operation. The coefficient of variation (CV) is the ratio of the standard deviation to the mean and shows the extent of variability in relation to the mean.
- F. **Formula to calculate TEA:** $\%TEA = \%bias + 2CV$

III. RECOMMENDED USE OF TEA CRITERIA:

- A. Evaluation criteria for lot-to-lot comparison using patient samples. Difference between lots generally should not exceed TEA.
- B. Evaluation criteria for method validation and verification. Bias between methods generally should not exceed 1/2 of TEA. Imprecision generally should not exceed 1/2 of TEA.
- C. The target CV for QC ranges generally should not exceed 1/2 of TEA.

Attachments

[Total Allowable Error \(TEA\) Reference Guide rev 1.2024](#)

Approval Signatures

Step Description	Approver	Date
Medical Directors	Vaishali Pansare: Chief, Pathology	2/6/2024
Medical Directors	Ann Marie Blenc: System Med Dir, Hematopath	2/1/2024
Medical Directors	Muhammad Arshad: Chief, Pathology	1/30/2024
Medical Directors	Jeremy Powers: Chief, Pathology	1/26/2024
Medical Directors	Ryan Johnson: OUWB Clinical Faculty	1/26/2024
Medical Directors	John Pui: Chief, Pathology	1/26/2024
Policy and Forms Steering Committee Approval (if needed)	Kelly Walewski: Supv, Laboratory	1/26/2024

Caitlin Schein: Staff Physician	1/25/2024
Qian Sun: Tech Dir, Clin Chemistry, Path	1/22/2024
Nga Yeung Tang: Tech Dir, Clin Chemistry, Path	1/16/2024
Jennifer Yaker: Mgr, Laboratory	1/16/2024
Kristin Russell: Supv, Laboratory	1/4/2024
Kristen DiCicco: Mgr, Laboratory	1/4/2024
Christopher Ferguson: Mgr, Laboratory	1/4/2024
Katherine Persinger: Mgr, Laboratory	1/4/2024
Ashley Beesley: Mgr, Laboratory	1/4/2024
Michelle Alexander: Medical Technologist Lead	1/4/2024
Kelly Walewski: Supv, Laboratory	1/4/2024
Leah Korodan: Mgr, Division Laboratory	1/4/2024

Applicability

Dearborn, Farmington Hills, Grosse Pointe, Royal Oak, Taylor, Trenton, Troy, Wayne