



# STANTON TERRITORIAL HEALTH AUTHORITY

## Yellowknife, Northwest Territories

<b>TITLE: <math>\beta</math>-Lactamase (Cefinase)</b>	<b>Revision Date:</b> 07-March-2016	<b>Issue Date:</b> 07-March-2014
<b>Document Number: MIC50300</b>	<b>Status: <span style="color: red;">Approved</span></b>	
<b>Distribution: Microbiology Test Manual</b>	<b>Page: 1 of 5</b>	
<b>Approved by:</b> C. Case, Manager of Diagnostic Services	<b>Signed by:</b> <i>Cheryl Case</i>	

### PRINCIPLE:

Certain organisms have acquired the capacity to produce enzymes ( $\beta$ -lactamases) that inactivate  $\beta$ -lactam antibiotics. Several tests are available to detect the presence of  $\beta$ -Lactamase, by using different principles. The chromogenic method, as described below, is considered to be the preferred method.

The Cefinase disc is impregnated with the chromogenic cephalosporin, Nitrocefin. This compound exhibits a very rapid colour change from **yellow** to **red** as the amide bond in the  $\beta$ -lactam ring is hydrolyzed by a  $\beta$ -lactamase. When a bacterium produces this enzyme in significant quantities, the **yellow** coloured disc turns **red** in the areas where the isolate is smeared.

### PURPOSE:

To rapidly test for the production of  $\beta$ -lactamase in colonies of *Enterococcus faecalis*, *Neisseria gonorrhoeae*, *Staphylococcus aureus*, *Moraxella catarrhalis*, *Haemophilus influenzae* and anaerobic bacteria.

### SAMPLE INFORMATION:

<b>Type</b>	Well isolated colonies
-------------	------------------------

<b>NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.</b>	
<b>FILENAME: MIC50300Beta-LactamasePRO.doc</b>	<b>PRINT DATE: 7 March 2014</b>

<b>TITLE: <math>\beta</math>-Lactamase (Cefinase)</b>	<b>Revision Date:</b> 07-March-2016	<b>Issue Date:</b> 07-March-2014
<b>Document Number: MIC50300</b>	<b>Status: <span style="color: red;">Approved</span></b>	
<b>Distribution: Microbiology Test Manual</b>	<b>Page: 2 of 5</b>	

**REAGENTS and/or MEDIA:**

	Reagent #1	Reagent #2
<b>Type</b>	Cefinase discs	Sterile water
<b>Volume</b>	One disc per bacterial strain	1 drop/disc
<b>Stability</b>	Do not use past expiry date	
<b>Storage Requirements</b>	-20°C-8°C in air tight container with dessicant	

**SUPPLIES:**

- Wooden sticks

**SPECIAL SAFETY PRECAUTIONS:**

Since viable microorganisms are used, all cultures must be handled with appropriate precautions. All equipment in contact with cultures should be contaminated by appropriate methods. Avoid contact of the cefinase discs with eyes or skin and refer to MSDS sheet for cefinase disc prior to use.

**QUALITY CONTROL:**

Quality control is set up each day the test is performed using the following control organisms. A QC order is generated in the TQC system.

**Positive:**     *Staphylococcus aureus*     ATCC 29213

**Negative:**    *Haemophilus influenzae*    ATCC 10211

<b>NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.</b>	
<b>FILENAME: MIC50300Beta-LactamasePRO.doc</b>	<b>PRINT DATE: 7 March 2014</b>

<b>TITLE: <math>\beta</math>-Lactamase (Cefinase)</b>	<b>Revision Date:</b> 07-March-2016	<b>Issue Date:</b> 07-March-2014
<b>Document Number: MIC50300</b>	<b>Status: <span style="color: red;">Approved</span></b>	
<b>Distribution: Microbiology Test Manual</b>	<b>Page: 3 of 5</b>	

### PROCEDURE INSTRUCTIONS:

Step	Action
<b>Testing a Pure Culture for <math>\beta</math>-lactamase Production</b>	
<b>1</b>	To order a $\beta$ -lactamase in the system, order <b>LIS CODE: ^BL</b>
<b>2</b>	Using forceps, dispense the required number of discs from the cartridge into an empty petri dish or onto a microscope slide.
<b>3</b>	Moisten each disc with 1 drop of sterile distilled water – do NOT flood the disc with water.
<b>4</b>	With a sterilized loop or applicator stick, remove several well-isolated similar colonies and smear onto a disc surface.
<b>5</b>	Observe disc for colour change within appropriate time-frame for organism tested. See chart under “Result Interpretation”.
<b>Alternate Procedure</b>	
<b>1</b>	Using forceps, moisten disc with one drop of sterile water and then wipe across colony.

### LIMITATIONS OF THE PROCEDURE:

Primary isolation medium (e.g. modified Thayer-Martin) may grow  $\beta$ -lactamase producing bacteria in addition to the pathogen, which may result in a false positive test. Therefore, ensure that the culture being tested is pure.

A negative  $\beta$ -lactamase result **does not imply** sensitivity to Penicillin or Ampicillin.

### INTERPRETATION OF RESULTS:

<b>POSITIVE</b>	<b>Red</b> colour development at site of inoculation
<b>NEGATIVE</b>	No colour change

<b>NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.</b>	
<b>FILENAME: MIC50300Beta-LactamasePRO.doc</b>	<b>PRINT DATE: 7 March 2014</b>

<b>TITLE: <math>\beta</math>-Lactamase (Cefinase)</b>	<b>Revision Date:</b> 07-March-2016	<b>Issue Date:</b> 07-March-2014
<b>Document Number: MIC50300</b>	<b>Status: <span style="color: red;">Approved</span></b>	
<b>Distribution: Microbiology Test Manual</b>	<b>Page: 4 of 5</b>	



For most bacterial strains a positive result will develop within 5 minutes. However, positive reactions for some staphylococci may take up to 1 hour to develop. Refer to the following chart for expected reaction times.

Organism	Result	Approx. Reaction Time	Interpretation
<i>Staphylococcus aureus</i>	Positive	1 hour	Resistant to penicillin, ampicillin, carbenicillin, and ticarcillin. Probably susceptible to cephalothin, methicillin, oxacillin, nafcillin and other penicillinase-resistant penicillins
<i>Enterococcus faecalis</i>	Positive	5 minutes	Resistant to penicillin and ampicillin
<i>Haemophilus influenzae</i>	Positive	1 minute	Resistant to ampicillin. Susceptible to cephalosporins.
<i>Neisseria gonorrhoeae</i> and <i>Moraxella catarrhalis</i>	Positive	1 minute	Resistant to penicillin
Anaerobic bacteria	Positive	30 minutes	Probable identification is <i>Bacteroides</i> species. Probably resistant to penicillin and may be resistant to cephalosporins including cefotaxime and rarely cefoxitin.

## REFERENCES:

- BBL, B. (2004/06). Paper Disc for the Detection of  $\beta$ -Lactamase Enzymes.  
*Package Insert*

<b>TITLE: <math>\beta</math>-Lactamase (Cefinase)</b>	<b>Revision Date:</b> 07-March-2016	<b>Issue Date:</b> 07-March-2014
<b>Document Number: MIC50300</b>	<b>Status: <span style="color: red;">Approved</span></b>	
<b>Distribution: Microbiology Test Manual</b>	<b>Page: 5 of 5</b>	

**REVISION HISTORY:**

<b>REVISION</b>	<b>DATE</b>	<b>Description of Change</b>	<b>REQUESTED BY</b>
1.0	03Nov10	Initial Release	ML Dufresne
1.1	31Jul13	Illustrations added	A Darrach
1.2	07Mar14	Changed to document control number MIC50300	C Russell

**NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.**

**FILENAME: MIC50300Beta-LactamasePRO.doc**

**PRINT DATE: 7 March 2014**