






 <p>1. From Home Screen, Press Login.</p>	 <p>2. Enter or scan Operator ID and press Accept.</p>	 <p>3. From Patient Test screen press QC.</p>
 <p>4. Enter or scan Strip lot no. and press Accept.</p>	 <p>5. Scan QC lot no. and press Accept.</p>	 <p>6. Insert Test Strip into Meter.</p>
 <p>7. Touch drop from QC bottle to strip.</p>	<p>8. WARNING!</p> <p>The test strip must fill completely upon touching the QC droplet. DO NOT add a second QC droplet to the test strip. Discard the test strip and repeat the test with a new test strip.</p>	 <p>9. To accept result press Accept.</p>

Test Strips have an open expiration of **180** days

Control Solutions have an open expiration of **90** days

ALL test strip bottles and control solutions **MUST** be labeled with the appropriate discard date

NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME: POC20102NovaBiomedicalGlucometerQuickQCGuideAID.doc

PRINT DATE: 7 July 2014