

Document Name: Gram stain resulting in LIS –
Sterile Fluids

Approved By:

Status: **DRAFT**

PURPOSE: To report the Gram stain results of STAT sterile fluid specimens in a consistent manner.

SAMPLE INFORMATION:

Type	<ul style="list-style-type: none"> • Sterile fluids, including CSF. • Refer to MIC10230 – Microbiology Specimen Processing.
-------------	-----------------------------------------------------------------------------------------------------------------------------------------------------

REAGENTS INFORMATION:

Type	BD™ Gram Crystal Violet, 3.8 L, B4312526 BD™ Gram Iodine (Stabilized), 3.8 L, B4312543 BD™ Gram Decolorizer, 3.8 L, B4312528 BD™ Gram Safranin, 3.8 L, B4312531
Source	Fisher Scientific Canada
Storage	Store at 15° to 30°
Stability	As per expiry date listed on bottle

SUPPLIES:

- Ringed cytology microscope slide
- Frosted end glass microscope slide
- QC slide
- Methanol, absolute
- Immersion oil
- Microscope
- Slide storage tray

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME:

Print Date:

Document Name: Gram stain resulting in LIS – Sterile Fluids	Document Number: MIC20400	
	Version No: 1.0	Page: 2 of 6
	Effective: DRAFT	

SPECIAL SAFETY PRECAUTIONS:

Containment Level 2 facilities, equipment, and operational practices for work involving infectious or potential infectious materials or cultures.

- Lab gown must be worn when performing activities with potential pathogens.
- Gloves must be worn when direct skin contact with infected materials is unavoidable.
- Eye protection must be used when there is a known or potential risk of exposure of splashes.
- All procedures that may produce aerosols, or involve high concentrations or large volumes should be conducted in a biological safety cabinet (BSC).
- The use of needles, syringes and other sharp objects should be strictly limited.


All patient specimens are assumed to be potentially infectious. Universal precautions must be followed. Since viable micro-organisms are used, all cultures must be handled with appropriate precautions. All equipment in contact with cultures should be decontaminated by appropriate methods.

QUALITY CONTROL:

- Refer to MIC60060 – Microbiology Stain Quality Control.

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.	
FILENAME:	Print Date:

PROCEDURE INSTRUCTIONS:

Step	Action
1	<p>Pending Gram stain orders are found in the LIS Resulting Worklist: Resulting Worklist → GRM/MISC_1of2</p>  <p>Press enter or double click to open worklist.</p>
2	<p>Enter the accession number on the slide and select enter to mark the order. Select enter again to open Result Entry or double click on accession number to open.</p>
3	<p>Under low power (10X/ LPF), screen slide to locate good specimen areas to obtain an overall impression of cell types present.</p> <ul style="list-style-type: none"> • Observe slide for stain crystals: <ul style="list-style-type: none"> ➤ If an excess of precipitated stain is observed, prepare another smear. ➤ If precipitate continues, use freshly filtered crystal violet. • Determine if slide has been properly decolorized: <ul style="list-style-type: none"> ➤ Depending on the source of the specimen, the background should be generally clear or Gram negative. ➤ If WBC are present, they should appear completely Gram negative. ➤ If slide is over decolorized, prepare another smear. • Determine if thickness of smear is appropriate: <ul style="list-style-type: none"> ➤ For proper interpretation, areas must be no more than one cell thick, with no overlapping of cells. Prepare a new slide if unreadable. • Examine for evidence of inflammation: <ul style="list-style-type: none"> ➤ Determine areas representative of inflammation and areas of contamination with squamous epithelial cells.

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.




4	Under the test code: STGM1 , use the STGM1 keypad to report the quantity of epithelial cells, white blood cells, red blood cells and bacteria seen. Report cells in this order to maintain consistency with reporting. Scan approximately 20 to 40 fields.										
5	<p>Epithelial cells, white blood cells and red blood cells are quantified as follows under LPF (10X):</p> <table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <td style="text-align: center;">None seen</td> <td style="text-align: center;">No cells seen</td> </tr> <tr> <td style="text-align: center;">1+</td> <td style="text-align: center;">< 1 cells seen</td> </tr> <tr> <td style="text-align: center;">2+</td> <td style="text-align: center;">1 - 9 cells seen</td> </tr> <tr> <td style="text-align: center;">3+</td> <td style="text-align: center;">10 - 25 cells seen</td> </tr> <tr> <td style="text-align: center;">4+</td> <td style="text-align: center;">> 25 cells seen</td> </tr> </table> <p>NOTE: Only report “None seen” for white blood cells. If no epithelial cells or red blood cells are seen, do not report this.</p>	None seen	No cells seen	1+	< 1 cells seen	2+	1 - 9 cells seen	3+	10 - 25 cells seen	4+	> 25 cells seen
None seen	No cells seen										
1+	< 1 cells seen										
2+	1 - 9 cells seen										
3+	10 - 25 cells seen										
4+	> 25 cells seen										
6	Add one drop of immersion oil to the slide. In a representative area with predominance of inflammation or purulence using the oil immersion lens (100X), examine 20 to 40 fields to observe cell morphology and Gram reaction.										
7	<p>Bacterial and yeast cells are quantified as follows under OIF (100x):</p> <table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <td style="text-align: center;">None seen</td> <td style="text-align: center;">No cells seen</td> </tr> <tr> <td style="text-align: center;">1+</td> <td style="text-align: center;">< 1 cells seen</td> </tr> <tr> <td style="text-align: center;">2+</td> <td style="text-align: center;">1 - 9 cells seen</td> </tr> <tr> <td style="text-align: center;">3+</td> <td style="text-align: center;">10 - 25 cells seen</td> </tr> <tr> <td style="text-align: center;">4+</td> <td style="text-align: center;">> 25 cells seen</td> </tr> </table> <p>NOTE: If no bacteria are seen, report this result.</p>	None seen	No cells seen	1+	< 1 cells seen	2+	1 - 9 cells seen	3+	10 - 25 cells seen	4+	> 25 cells seen
None seen	No cells seen										
1+	< 1 cells seen										
2+	1 - 9 cells seen										
3+	10 - 25 cells seen										
4+	> 25 cells seen										
8	Be wary of interpretations made from observing very few organisms (especially in the absence of inflammation or if the organisms are unevenly distributed), as collection tubes, slides and media may harbor nonviable bacteria. For sterile fluids, where the results will define an infectious process, prepare a second smear to confirm rare findings of microorganisms.										
9	Finalize STGM1. Preview instant report and save. Refresh GRM/MISC1of2 worklist. If finished reading slides, ensure Gram stains remaining on worklist have been prepared to be read at a later time.										
10	Gently blot excess oil from slide using paper towel or gauze and save slides for further evaluation on the slide tray designated for day slides being read.										

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME:

Print Date:

REPORTING OF RESULTS:

IF	REPORT
No white blood cells seen on Gram stain	Report: “No white blood cells seen”
No bacteria seen on Gram stain	Report: “No bacteria seen”
Epithelial cells, white blood cells and red blood cells seen on Gram stain	<ul style="list-style-type: none"> Quantitate and report using the STGM1 keypad.
Bacteria seen on Gram stain	<ul style="list-style-type: none"> Quantitate and report using the STGM1 keypad. Bacteria seen in the Gram stain of sterile fluids are considered a critical result. Phone ordering location to give result. Document call in the “Call” box. Refer to LIS?????. If unable to reach ordering location, consult the hospital wide policy “Laboratory: Critical Values – Responsible Party”.
Bacteria resembles: <i>Staphylococcus spp.</i> 	Report: “Gram positive cocci suggestive of Staphylococci” Note: Use caution if in doubt. If doubt exists, report as Gram-positive cocci.
Bacteria resembles: <i>Streptococcus spp.</i> 	Report: “Gram positive cocci suggestive of Streptococci” Note: Use caution if in doubt. If doubt exists, report as Gram-positive cocci.
Bacteria resembles: Diphtheroids 	Report: “Gram positive bacilli resembling diphtheroids” Note: Use caution if in doubt. If doubt exists, report as Gram-positive bacilli.

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME:

Print Date:

LIMITATIONS:

1. If rare or no organisms are seen from a normally sterile site but the specimen appears purulent, or the specimen looks suspicious, perform more extensive review of the slide.
2. Critical specimens where low numbers of organisms may be seen (e.g.; CSF) should be read more extensively and the entire slide may need to be read to substantiate a negative result.
3. The presence of a microorganism from a normally sterile site is likely to indicate infection with that organism.

REFERENCES:

- Clinical Microbiology Procedures Handbook, 4th edition, ASM Press, 2016.

REVISION HISTORY:

REVISION	DATE	Description of Change	REQUESTED BY
1.0		Initial Release	L. Steven