



Stanton Territorial Hospital
P.O. Box 10, 550 Byrne Road
YELLOWKNIFE NT X1A 2N1

Document Name:
Vitek 2 Quality Control Job Aid

Distribution:
Microbiology Quality Control Manual

Document Number:
MIC60031

Date Issued:
DRAFT

Weekly/New Lot Number/New Shipment Vitek QC:

QC Organism	ATCC #	Vitek Card
<i>Staphylococcus aureus</i>	29213	AST-GP67
<i>Staphylococcus aureus</i>	BAA-1026	AST-GP67
<i>Staphylococcus aureus</i>	BAA-977	AST-GP67
<i>Enterococcus faecalis</i>	29212	AST-GP67
<i>Enterococcus faecalis</i>	51299	AST-GP67
<i>Escherichia coli</i>	25922	AST-N390
<i>Escherichia coli</i>	35218	AST-N390
<i>Pseudomonas aeruginosa</i>	27853	AST-N390

New Lot Number/New Shipment QC:

QC Organism	ATCC #	Vitek Card
<i>Enterococcus casseliflavus</i>	700327	GP
<i>Staphylococcus saprophyticus</i>	BAA-750	GP

QC Organism	ATCC #	Vitek Card
<i>Enterobacter hormaechei</i>	700323	GN
<i>Stenotrophomonas maltophilia</i>	17666	GN

QC Organism	ATCC #	Vitek Card
<i>Candida albicans</i>	14053	YST
<i>Candida glabrata</i>	MYA-2950	YST

QC Organism	ATCC #	Vitek Card
<i>Eikenella corrodens</i>	BAA-1152	NH

NOTE: This is a controlled document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME:

Print Date: