

PROGRAM Standard Operating Procedure – Laboratory Services	
Title: MIC32400 – Ear Culture	Policy Number:
Program Name: Laboratory Services	
Applicable Domain: Lab, DI and Pharmacy Services	
Additional Domain(s):	
Effective Date:	Next Review Date:
Issuing Authority: Director of Health Services	Date Approved:
Accreditation Canada Applicable Standard: N/A	

**GUIDING PRINCIPLE:**

Otitis media is an infection of the middle ear. Complications include hearing loss and learning difficulties. Common causes of otitis media are *Streptococcus pneumoniae*, *Haemophilus influenzae*, *Moraxella catarrhalis* and *Alloiococcus otitis*. *Streptococcus pyogenes* is found on a seasonal basis. Treatment failure can occur and surgical intervention may be necessary. Otitis externa is an infection of the external auditory canal. This narrow canal tends to trap foreign objects, wax and water. Infections are classified as acute and chronic. Acute infections (often called swimmer’s ear) are frequently caused by *Pseudomonas aeruginosa*, although other aerobic organisms can be involved. *Vibrio alginolyticus* is a cause in oceanic swimmers. Localized infections with *Staphylococcus aureus* or *Streptococcus pyogenes* can also occur. Contaminating skin microbiota such as coryneform organisms (diphtheroids) and coagulase-negative *Staphylococci* may be present but are not significant. Chronic infections are usually caused by fungi, Mycobacteria, *Nocardia* and underlying diseases such as syphilis.

**PURPOSE/RATIONALE:**

To determine the presence or absence of bacterial pathogens in ear specimens.

**SCOPE/APPLICABILITY:**

This procedure applies to Medical Laboratory Technologists (MLTs) processing specimens for ear culture.

### SAMPLE INFORMATION:

<b>Type</b>	Swab <ul style="list-style-type: none"><li>• Amie's with or with charcoal</li></ul>
<b>Source</b>	<ul style="list-style-type: none"><li>• External auditory canal (outer ear)</li><li>• Otitis media discharge swabbed from external auditory canal</li></ul>
<b>Stability</b>	If the sample is received in the laboratory and processed greater than 48 hours from collection: <ul style="list-style-type: none"><li>• Add specimen quality comment: "Delayed transport may adversely affect pathogen recovery"</li></ul>
<b>Storage Requirements</b>	Room temperature. If transport is > 2 hours, swabs should be refrigerated.
<b>Criteria for rejection</b>	<ol style="list-style-type: none"><li>1. Unlabeled/mislabeled swabs</li><li>2. Specimen container label does not match patient identification on requisition</li></ol>

### NOTE:

- Refer to MIC34100-Body Fluid Culture for typanocentesis fluid

### REAGENTS and/or MEDIA:

- Blood agar (BA), Chocolate agar (CHO) and MacConkey agar (MAC)
- Identification reagents: catalase, oxidase, Staph latex test, Strep latex test, etc.

### SUPPLIES:

- Disposable inoculation needles
- Microscope slides
- Wooden sticks

### EQUIPMENT

- Biosafety cabinet
- 35° ambient air and 35° CO<sub>2</sub> incubators
- Vitek 2 and supplies

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

### SPECIAL SAFETY PRECAUTIONS:

Containment Level 2 facilities, equipment, and operational practices for work involving infectious or potential infectious materials or cultures.

- Ensure that appropriate hand hygiene practices be used.
- Lab gown must be worn when performing activities with potential pathogens.
- Gloves must be worn when direct skin contact with infected materials is unavoidable.
- Eye protection must be used when there is a known or potential risk of exposure of splashes.
- All procedures that may produce aerosols, or involve high concentrations or large volumes should be conducted in a biological safety cabinet (BSC).
- The use of needles, syringes and other sharp objects should be strictly limited.

All patient specimens are assumed to be potentially infectious. Routine Practices must be followed. Since viable micro-organisms are used, all cultures must be handled with appropriate precautions. All equipment in contact with cultures should be decontaminated by appropriate methods.

### QUALITY CONTROL:

- Refer to Test Manual for reagent quality control procedures

### PROCEDURE INSTRUCTIONS:

Step	Action
<b>Processing specimens for ear culture</b>	
1	In the biosafety cabinet: <ul style="list-style-type: none"><li>• Inoculate BA, CHO and MAC with the swab</li><li>• Ensure all surfaces of swab make contact with the agar</li><li>• Streak for isolated growth using a disposable inoculation needle</li><li>• Prepare smear by rolling the swab gently across the slide to avoid destruction of cellular elements and disruption of bacterial arrangements</li></ul>
2	Incubate the media: <ul style="list-style-type: none"><li>• Place BA and CHO in the CO<sub>2</sub> incubator</li><li>• Place MAC in the O<sub>2</sub> incubator</li></ul>
3	Allow smear to dry and perform gram stain. Gram stain must be read before culture plates. Refer to MIC20115-Gram stain procedure.

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

Probable Pathogens	Commensal Flora
<ul style="list-style-type: none"> <li>• <i>Staphylococcus aureus</i></li> <li>• <i>Streptococcus pyogenes</i></li> <li>• Other <math>\beta</math>-hemolytic <i>Streptococci</i></li> <li>• <i>Streptococcus pneumoniae</i></li> <li>• <i>Haemophilus influenzae</i></li> <li>• <i>Moraxella catarrhalis</i></li> <li>• <i>Pseudomonas aeruginosa</i></li> <li>• Non-fermentative Gram-negative bacilli</li> <li>• Enterobacteriaceae spp.</li> <li>• <i>Candida</i> spp.</li> <li>• Fungi</li> </ul>	<ul style="list-style-type: none"> <li>• <i>Aerococcus</i> spp.</li> <li>• <i>Bacillus</i> spp.</li> <li>• Coagulase negative <i>Staphylococcus</i></li> <li>• <i>Corynebacterium</i> spp.</li> <li>• <i>Micrococcus</i> spp.</li> <li>• <i>Neisseria</i> spp.</li> <li>• Anaerobes</li> <li>• viridans <i>Streptococcus</i> grp.</li> </ul>

### INTERPRETATION OF RESULTS:

Step	Action
<b>1</b>	Ensure growth on culture media correlates with gram stain results. If discordant results are found between the gram stain and growth: <ul style="list-style-type: none"> <li>• Re-examine smear and culture plates</li> <li>• Check for anaerobic growth</li> <li>• Re-incubate media to resolve</li> <li>• Consider re-smearing or re-planting specimen</li> </ul>
<b>2</b>	<ul style="list-style-type: none"> <li>• Observe BA and CHO plates at 24 hours and 48 hours</li> <li>• Observe MAC plate at 24 hours</li> <li>• Examine for growth of enteric Gram-negative bacilli, <i>Pseudomonas aeruginosa</i>, Yeasts, <i>Vibrio alginolyticus</i>, <math>\beta</math>-hemolytic <i>Streptococci</i>, <i>Staphylococcus aureus</i>, <i>Haemophilus influenzae</i>, <i>Moraxella catarrhalis</i> and Fungi.</li> </ul>
<b>3</b>	If specimen is discharge from otitis media infection, the most common pathogens are <i>Streptococcus pyogenes</i> , <i>Streptococcus pneumoniae</i> , <i>Moraxella catarrhalis</i> and <i>Haemophilus influenzae</i> .
<b>4</b>	Skin flora such as coagulase negative <i>staphylococci</i> and coryneforms are normal in the external ear canal and should not be further evaluated.
<b>5</b>	Mixed cultures of Gram-negative rods should be minimally identified.

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

**REPORTING INSTRUCTIONS:**

IF	REPORT
No growth after 1 day	<b>PRELIM:</b> <ul style="list-style-type: none"> <li>Report: <b>"No Growth after 1 Day. Further report to follow"</b></li> </ul>
No growth after 2 days	<b>FINAL:</b> <ul style="list-style-type: none"> <li>Report: <b>"No Growth after 2 Days"</b></li> </ul>
Mix of skin flora	<ul style="list-style-type: none"> <li>Report: <b>"Mixture of skin flora"</b></li> <li>List quantitation</li> </ul>
Mix of enteric Gram-negative bacilli	<ul style="list-style-type: none"> <li>Report <b>"Mixture of coliform organisms"</b></li> <li>List quantitation</li> </ul>
Growth or mix of other non-pathogenic organisms	<ul style="list-style-type: none"> <li>Report <b>"Commensal flora"</b> or <b>"Commensal skin flora"</b></li> <li>List quantitation</li> </ul>
Growth of pathogen(s)	<ul style="list-style-type: none"> <li>Report organisms(s) identification</li> <li>List quantitation</li> <li>Report susceptibility results as per ASTM</li> </ul>

**NOTE:**

- Refer to Reportable Diseases – Public Health Act as of September 2009 for reporting to HPU1
- Refer to MIC35100-Nosocomial Infection Notification Job Aid to determine if organism needs to be copied to Stanton Infection Prevention and Control
- Refer to L-0910-Laboratory: Critical Values for results that need to be phoned to ordering location
- Refer to MIC10510-Referral of Category B Specimens to *DynaLIFE* and Alberta Precision Laboratories for sending isolates to *DynaLIFE*

**LIMITATIONS:**

- An external ear culture with a predominant Gram-negative bacillus, beta-hemolytic *Streptococcus*, *Staphylococcus aureus*, *Streptococcus pneumoniae*, *Haemophilus influenzae* or *Moraxella catarrhalis* generally indicates infection with that organism.
- Historically, the nasopharynx was cultured in order to predict the pathogens in the middle ear. This practice is no longer recommended.
- Treatment of otitis externa usually consists of ear drops.
- False-negative cultures can result from overgrowth of the culture with normal skin flora.
- False-positive results can be caused by over-interpretation of culture results.

**CROSS-REFERENCES:**

- MIC10510-Referral of Category B Specimens to *DynaLIFE* and Alberta Precision Laboratories
- MIC20115-Gram stain procedure
- MIC34100-Body Fluid Culture for typanocentesis fluid
- MIC35100-Nosocomial Infection Notification Job Aid

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

**REFERENCES:**

1. Leber, A. (2016). *Clinical microbiology procedures handbook*. (4<sup>th</sup> ed.) Washington, D.C.: ASM Press
2. Jorgensen J.H., Pfaller M.A., Carroll K.C., Funke G., Landry M.L., Richter S.S., Warnock D.W. (2015). *Manual of Clinical Microbiology, 11<sup>th</sup> edition*. Washington, D.C: ASM Press

**APPROVAL:**

\_\_\_\_\_  
Date

**REVISION HISTORY:**

REVISION	DATE	Description of Change	REQUESTED BY
1.0	06 Nov 17	Initial Release	L. Steven
2.0	04 Dec 18	Updated to include new Vitek 2 instrument	L. Steven
3.0	26 Feb 21	Procedure reviewed and added to NTHSSA policy template	L. Steven

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.