



**Stanton Territorial Hospital**  
 P.O. Box 10, 550 Byrne Road  
 YELLOWKNIFE NT X1A 2N1

**Document Name:**  
 Maintenance Record – BACTEC FX

**Distribution:**  
 Microbiology Instrumentation Manual

**Document Number:**  
 MIC711110

**Date Issued:**  
 10 May 2019

**Maintenance Record – BACTEC FX**

**Month:** \_\_\_\_\_ **Year:** \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
---	---	---	---	---	---	---	---	---	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----

**Daily Maintenance – Left Side (A)**

Temp: 33.5-36.5																																
Internal temp																																
Green LEDs																																
Red LEDs																																
Yellow LEDs																																

**Daily Maintenance – Right Side (B)**

Temp: 33.5-36.5																																
Internal temp																																
Green LEDs																																
Red LEDs																																
Yellow LEDs																																
Initials																																

**Monthly Maintenance**

Change air filter	Date	Tech	Change air filter	Date	Tech
Black air filter			Silver air filter		

Corrective action: \_\_\_\_\_ Reviewed by: \_\_\_\_\_

**NOTE:** This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.



**Stanton Territorial Hospital**  
P.O. Box 10, 550 Byrne Road  
YELLOWKNIFE NT X1A 2N1

**Document Name:**  
Maintenance Record – BACTEC FX

**Distribution:**  
Microbiology Instrumentation Manual

**Document Number:**  
MIC711110

**Date Issued:**  
10 May 2019

**NOTE:** This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME: MIC71110MaintenanceRecord-BACTECFXFOR.doc

PRINT DATE: 6/22/2021 12:20:00 PM