



**Stanton Territorial Hospital**  
P.O. Box 10, 550 Byrne Road  
YELLOWKNIFE NT X1A 2N1

**Document Name:**  
Vitek 2 Card Job Aid

**Distribution:**  
Microbiology Instrumentation Manual

**Document Number:**  
MIC70420

**Date Issued:**  
3 June 2021

Card	McFarland Std.	Notes
GNI	0.48 to 0.63	
GNS	0.48 to 0.63	<ul style="list-style-type: none"><li>If mated GNI card not run, need to enter organism identification</li></ul>
GPI	0.48 to 0.63	
GPS	0.48 to 0.63	<ul style="list-style-type: none"><li>If mated GPI card not run, need to enter organism identification</li></ul>
YST	1.80 to 2.20	
NH	2.70 to 3.30	
ANC	2.70 to 3.30	<ul style="list-style-type: none"><li>Need to enter offline tests: gram stain, morphology and aerotolerance</li></ul>

**NOTE:** This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME: MIC70420-Vitek 2 Card Job Aid.doc.docx

PRINT DATE: 6/14/2021 12:10:00 PM