



Stanton Territorial Hospital
 P.O. Box 10, 550 Byrne Road
 YELLOWKNIFE NT X1A 2N1

Document Name:
 Maintenance Record – BioFire

Distribution:
 Microbiology PCR Testing Manual

Document Number:
 MIC82110

Date Issued:
 22 Apr 2020

Maintenance Record – BioFire

Month/Year: _____ **Instrument:** _____

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
---	---	---	---	---	---	---	---	---	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----

Daily Maintenance

Disinfect work area																															
Check printer paper																															

Monthly Maintenance

Disinfect Torch top and sides																														
Disinfect Torch screen																														
Disinfect Torch module covers																														
Change fan filters																														
Clean used fan filters																														
Archive tests																														

As-Required

Switching to admin mode																														
Decontamination after pouch leakage																														
Check function of decontaminated module																														

Corrective action: _____ Reviewed by: _____
