

| <b>PROGRAM Standard Operating Procedure – Laboratory Services</b> |                   |
|---|-------------------|
| Title: MIC60010 –<br>Microbiology Quality Control                 | Policy Number:    |
| Program Name: Laboratory Services                                 |                   |
| Applicable Domain: Lab, DI and Pharmacy Services                  |                   |
| Additional Domain(s):   |                   |
| Effective Date:   | Next Review Date: |
| Issuing Authority:<br>Director, Health Services                   | Date Approved:    |
| Accreditation Canada Applicable Standard: N/A                     |                   |

**GUIDING PRINCIPLE:**

Quality control testing that is out of range will be repeated. If result continues out of range, it is considered a non-conformance. For all **non-conformances**:

- Take the non-conforming item out of service. Clearly label as unsatisfactory and place into segregation
- Withhold patient results obtained with the unacceptable material/equipment/instrument
- Contact the vendor for further instructions if required
- Notify patient's unit or clinician if results will be delayed and explain why
- Repeat patient testing once the material/equipment/instrument has been successfully quality controlled
- For non-conformances related to media, reagents, susceptibility testing and equipment, notify the Technical Supervisor, Microbiology
- For non-conformances related to instrumentation, notify the Technical Supervisor, Microbiology and document the non-conformance on QUA40590-Instrument Troubleshooting for the instrument involved

**PURPOSE/RATIONALE:**

To ensure that patient results are consistently of the highest quality, all aspects of specimen testing in the Microbiology Laboratory will be quality controlled.

**SCOPE/APPLICABILITY:**

This procedure applies to Medical Laboratory Technologists (MLTs) performing quality control in the microbiology laboratory.

### **Anaerobic jars and trays:**

- Include a chemical indicator (resazurin) in all anaerobic jars and trays
- Daily, enter indicator results into TQC for each jar or tray opened
  - Refer to MIC61030-Entering Microbiology QC Results into TQC
- For chemical QC failure (pink indicator):
  - Refer to **non-conformances** section on page 1
  - Do not report results of agar plates from QC failed jars and trays (except for throat cultures positive for Group A *Streptococcus* and blood cultures growing aerobic/facultative organisms that match direct Gram-stained smear findings)
  - Re-plant all other specimens from failed jars and trays

### **API 20E strips:**

- The strips and reagents are systematically quality controlled at various stages of the manufacturing process
- Test new shipments and new lot numbers of API 20E strips on receipt using *Proteus mirabilis* ATCC 35659
- A QC order will generate in TQC when a new kit is received
- Enter results into TQC. Refer to MIC61030-Entering Microbiology QC Results into TQC
- If quality control testing results are acceptable and kit is a new lot number, place yellow "NEW LOT Record #" sticker on the first package (of same lot number) to be used
- If QC testing results are not acceptable, refer to **non-conformances** section on page 1. Repeat quality control testing with new/replacement kit
- The package inserts for API 20E strips and reagents are available online and in the "Package Inserts" binder in the Dark Room

### **API NH strips:**

- The strips and reagents are systematically quality controlled at various stages of the manufacturing process
- Test new shipments and new lot numbers of API NH strips on receipt using *Neisseria gonorrhoeae* ATCC 31426
- A QC order will generate in TQC when a new kit is received
- Enter results into TQC. Refer to MIC61030-Entering Microbiology QC Results into TQC
- If quality control testing results are acceptable and kit is a new lot number, place yellow "NEW LOT Record #" sticker on the first package (of same lot number) to be used
- If QC testing results are not acceptable, refer to **non-conformances** section on page 1. Repeat quality control testing with new/replacement kit
- Change reconstituted API NH James reagent monthly and ZYM B reagent every 2 weeks. Write the expiry date on reagent bottles
- If reagent is a new lot number, activate in TQC. Refer to MIC61020-Opening and Closing Lot Numbers in TQC
- The package inserts for API NH strips and reagents are available online and in the "Package Inserts" binder in the Dark Room

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

**BACTEC FX blood culture instrument:**

- Perform daily and monthly maintenance on the BACTEC FX instrument. Record on MIC71110-Maintenance Record-BACTEC FX
- Annual preventative maintenance is performed by BD. File records in the "BACTEC FX Service Reports and Error Log" binder in the Dark Room
- Record all non-conformances and issues, including UPS issues, with actions taken and resolutions on QUA40590-Instrument Troubleshooting in the "BACTEC FX Service Reports and Error Log" binder and notify the Technical Supervisor, Microbiology

**BioFire instrument:**

- Perform daily and monthly maintenance on both BioFire bases. Record on MIC82110-Maintenance Record-BioFire
- Record all non-conformances and issues with actions taken and resolutions on QUA40590-Instrument Troubleshooting in the "BioFire Service Reports and Error Log" binder and notify the Technical Supervisor, Microbiology
- Refer to specific test quality control procedures for quality control testing

**Biological safety cabinets:**

- Perform daily, weekly and bi-annual maintenance as required. Refer to MIC61100-Microbiology Laboratory Equipment
- BSCs are inspected, tested, and certified annually

**Disinfection and decontamination procedures:**

- Disinfect counter tops with Accel TB wipes at the end of the day and as required
- Dispose of sharps (needles, glass slides, glass tubes, broken glass) in yellow plastic sharps container. When 2/3 full, close container and place outside the microbiology laboratory in waste pickup area
- Dispose of contaminated materials such as culture plates, Vitek cards, swabs, sticks and specimens in large cardboard biohazard waste containers. When container is full and weighs no more than 16 kg, seal plastic biohazard waste bag and close up cardboard box. Place box outside the microbiology laboratory in waste pickup area
- Wipe the specimen processing cart and specimen buckets with Accel TB wipes monthly
- For biological spills or contamination, refer to the Code Brown procedure located in the code binder on the shared drive and on the laboratory safety wall

**External Quality Assurance Programs-CMPT, CAP and QASI:**

- The Microbiology Laboratory participates in external quality assessment programs (EQA) administered by Clinical Microbiology Proficiency Testing (CMPT), the College of American Pathologists (CAP) and (QASI)
- EQA specimens that are completed by MLTs will be documented on QUA30310-Laboratory Competence Assessment Form
- Treat specimens as much as possible as patient specimens

### **Freezer:**

- Daily, during morning start up, check digital temperature and internal thermometer temperature of the -80 freezer and record on temperature chart QUA40491.9-Temperature Log-Micro Ultra-Low Freezer
- If temperature is out of range, attempt to adjust. If temperature continues to be out of range:
  - Contact maintenance and notify the Technical Supervisor, Microbiology
  - Relocate all supplies from the faulty freezer to a functioning freezer operating at the appropriate temperature
  - Submit a RL6 incident report
  - After freezer has been serviced, document that the temperature has been within acceptable limits for 24 hours before placing back into use
- Perform weekly, monthly, and bi-annual maintenance as required
- Refer to MIC61100-Microbiology Laboratory Equipment
- Thermometers are replaced and verified annually and recorded on QUA40493-Thermometer Log

### **GeneXpert instrument:**

- Perform daily and monthly maintenance on both GeneXpert instruments. Record on MIC81110-Maintenance Record-GeneXpert
- Record all non-conformances and issues with actions taken and resolutions on QUA40590-Instrument Troubleshooting in the "GeneXpert Service Reports and Error Log" binder and notify the Technical Supervisor, Microbiology
- Refer to specific test quality control procedures for quality control testing

### **Incubators:**

- Daily, during morning start up, check digital temperature and internal thermometer temperature of all incubators. Record on temperature charts QUA40491.3, QUA40491.7 and QUA40491.8
- If temperature is out of range, attempt to adjust. If temperature continues to be out of range:
  - Contact maintenance and notify the Technical Supervisor, Microbiology
  - Relocate materials from the faulty incubator to a functioning incubator
  - Submit a RL6 incident report
  - After incubator has been serviced, document that the temperature has been within acceptable limits for 24 hours before placing back into use
- Perform weekly, monthly, and bi-annual maintenance as required. Refer to MIC61100-Microbiology Laboratory Equipment
- Thermometers are replaced and verified annually and recorded on QUA40493-Thermometer Log
- CO<sub>2</sub> incubators:
  - Check CO<sub>2</sub> gas gauge daily for adequate gas supply
  - Monthly measure the % CO<sub>2</sub> using the FYRITE gas analyzer. Record on QUA40491.7 and QUA40491.8
  - Result should be between 4-10%
  - If a CO<sub>2</sub> incubator alarms, use the manual stored on the Microbiology Shared Drive to try to troubleshoot the problem. If cannot solve issue, phone maintenance

- CO<sub>2</sub> tank change:
  - During regular hours, call Maintenance and request a new tank
  - Allow 1 hour for CO<sub>2</sub> incubator chamber to equilibrate. Ensure that the incubator door has not been opened for 30 minutes and perform FYRITE gas analyses on CO<sub>2</sub> incubators
  - Record result of FYRITE analyses on QUA40491.7 and QUA40491.8
  - Result should be between 4-10%

#### **Media quality control:**

- Quality control of prepared media is performed in accordance with CLSI M22-A3:2004 Quality Control for Commercially Prepared Microbiological Culture Media
- Refer to MIC61000-Receiving Supplies in the Microbiology Laboratory
- Refer to MIC60040-Culture Media Quality Control
- Store all media as specified by the manufacturer
- Certificate of Analysis are available online on the Oxoid website

#### **Microscopes:**

- Daily, check Kohler illumination
- All microscopes receive annual preventative maintenance
- Refer to MIC61100-Microbiology Laboratory Equipment for maintenance of microscopes and MIC61120-Kohler Illumination Job Aid

#### **Quality control records audit procedure:**

- The Technical Supervisor, Microbiology reviews completed quality control records in TQC on a weekly basis
- The Technical Supervisor, Microbiology reviews completed QC results records stored in accompanying QC results binder on a weekly basis
- Completed quality control records in TQC are kept indefinitely and completed manual QC results records are scanned onto the Microbiology Tech II shared drive

#### **Reagents:**

- Do not use expired reagents or solutions
- Perform quality control testing as per MIC60011-Microbiology QC Job Aid
- A QC order will generate in TQC when a new reagent is received, daily, weekly, or as tested depending on reagents QC requirements. Refer to MIC60011-Microbiology Quality Control Job Aid
- Enter results into TQC. Refer to MIC61030-Entering Microbiology QC Results into TQC
- If QC testing results are not acceptable, do not report patient results
- Repeat quality control testing. If repeat testing is acceptable, repeat patients test and report results. If repeat testing is not acceptable:
  - Refer to **non-conformances** section on page 1
  - Discard or return reagent to manufacturer as directed
  - Perform quality control testing on new reagent. If results are acceptable, repeat patients test and report results
- Package inserts are available online and in the "Product Inserts" binder

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.

### **Refrigerators:**

- Daily, during morning start up, check digital temperature and internal thermometer temperature of all refrigerators. Record on temperature charts QUA40491.1, QUA40494.2 and QUA40491.10
- If temperature is out of range, attempt to adjust. If temperature continues out of range:
  - Contact maintenance and notify the Technical Supervisor, Microbiology
  - Relocate all supplies from the faulty refrigerator to a functioning refrigerator operating at the appropriate temperature
  - Submit a RL6 incident report
  - After refrigerator has been serviced, document that the temperature has been within acceptable limits for 24 hours before placing back into use
- Perform weekly, monthly, and bi-annual maintenance as required. Refer to MIC61100-Microbiology Laboratory Equipment
- Thermometers are replaced and verified annually and recorded on QUA40493-Thermometer Log

### **Reporting results:**

- Patient reports are finalized in SoftMic by MLTs performing the specimen workup
- Notify the requester of the test when examinations are delayed and the turn-around time for a test is affected to a degree that may pose clinical implications
- Gram stain results must correlate with growth in culture
- Refer to MIC36000-Reportable and Communicable Diseases Notifications for notification of reportable and/or communicable diseases
- Refer to L-0910-Laboratory: Critical Values for notification of critical results
- Refer to MIC10210-Microbiology Rejection Criteria Job Aid, for specimen rejection criteria

### **Stock cultures:**

- Use only ATCC microorganisms for quality control testing
- A "Biosafety Risk Assessment Form for Organism Assessment" has been completed on each ATCC organism and is stored on the Biosafety and Biosecurity folder on the shared drive
- Store lyophilized cultures in the reagent refrigerator
- Store culture aliquots in glycerol citrate in the -70° freezer
- Refer to MIC60070-Stock Culture Maintenance for the handling of stock cultures; annually, monthly, and weekly
- Pathogen Safety Data Sheets (PSDSs) are technical documents produced by the Public Health Agency of Canada that describe the hazardous properties of human pathogens with recommendations for working with these agents. PSDSs are available at:  
<http://www.phac-aspc.gc.ca/lab-bio/res/psds-ftss/index-eng.php>

### **Susceptibility testing:**

- Store Active and Inactive E-test strips in the microbiology reagent refrigerator. Remove 1 hour before use
- Store Active and Inactive antimicrobial disks in the microbiology reagent refrigerator. Remove 1 hour before use
- Refer to MIC60020-Antibiotic Quality Control for QC procedure
- Perform quality control testing as directed on MIC60021-Antibiotic Quality Control Job Aid
- The weekly susceptibility QC order will generate in TQC every Wednesday
- Enter results into TQC. Refer to MIC61030-Entering Microbiology QC Results into TQC
- Verify any unusual results, repeat testing, and notify the Technical Supervisor, Microbiology
- For a list of unusual results, refer to current CLSI Performance Standards for Antimicrobial Susceptibility Testing M100-S, Appendix B-Intrinsic Resistance

### **Temperature and humidity monitoring:**

- Daily, during morning start up, check temperature and humidity of the Microbiology laboratory. Document on QUA40492.2-Room Temperature and Humidity Log
- Instrumentation acceptable ranges are:

|                     | Temperature  | Relative humidity |
|---------------------|--------------|-------------------|
| bioMerieux Vitek 2: | 15°C to 30°C | 20%-80%           |
| bioMerieux BioFire  | 15°C to 30°C | 20%-80%           |
| BD BACTEC FX        | 18°C to 30°C | 25%-80%           |
| Cepheid GeneXpert   | 15°C to 30°C | 10%-95%           |
- Notify the Technical Supervisor, Microbiology when readings are out of range
- Thermometers are replaced and verified annually and recorded on QUA40493-Thermometer Log

### **Vitek 2 identification and susceptibility testing system:**

- Perform daily, weekly, and monthly maintenance on the Vitek 2 instrument. Record on MIC70111-Maintenance Record-Vitek 2
- Refer to MIC60030-Vitek 2 Quality Control for QC procedure
- Perform quality control testing as directed on MIC60031-Vitek 2 Quality Control Job Aid
- Document quality control testing on MIC60032-QC Results Record-Vitek 2
- Record all non-conformances and issues, including UPS issues, with actions taken and resolutions on QUA40590-Instrument Troubleshooting in the "VITEK 2 Service Reports and Error Log" and notify the Technical Supervisor, Microbiology

### CROSS-REFERENCES:

- L-0910-Laboratory: Critical Values for notification of critical results
- QUA30310-Laboratory Competence Assessment Form
- QUA40493-Thermometer Log
- QUA40491.1,.2, .3, .5, .6, .7, .8, .9, .10
- QUA40590-Instrument Troubleshooting
- MIC10210-Microbiology Rejection Criteria Job Aid
- MIC36000-Notification of reportable and/or communicable diseases
- MIC60011-Microbiology QC Job Aid
- MIC60020-Antibiotic Quality Control
- MIC60021-Antibiotic Quality Control Job Aid
- MIC60030-Vitek 2 Quality Control
- MIC60031-Vitek 2 Quality Control Job Aid
- MIC60032-QC Results Record-Vitek 2
- MIC60040-Culture Media Quality Control
- MIC60070-Stock Culture Maintenance
- MIC61000-Receiving Supplies in the Microbiology Laboratory
- MIC61020-Opening and Closing Lot Numbers in TQC
- MIC61030-Entering Micro QC Results into TQC
- MIC61100-Microbiology Laboratory Equipment
- MIC70111-Maintenance Record-Vitek 2
- MIC71110-Maintenance Record BACTEC FX

### REFERENCES:

1. CLSI. *Quality Control for Commercially Prepared Microbiological Culture Media; Approved Standard—Third Edition*. CLSI document M22-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2004.
2. CLSI. *Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically*. 11th ed. CLSI standard M07. Wayne, PA: Clinical and Laboratory Standards Institute; 2018.
3. CLSI. *Performance Standards for Antimicrobial Disk Susceptibility Tests*. 13th ed. CLSI standard M02. Wayne, PA: Clinical and Laboratory Standards Institute; 2018.
4. CLSI. *Quality Control for Commercial Microbial Identification Systems; Approved Guideline*. CLSI document M50-A. Wayne, PA: Clinical and Laboratory Standards Institute; 2008.
5. Vitek 2 Instrument User Manual.
6. BACTEX FX Instrument User Manual.



**APPROVAL:**

\_\_\_\_\_  
Date

\_\_\_\_\_

**REVISION HISTORY:**

| REVISION | DATE      | Description of Change                                  | REQUESTED BY |
|----------|-----------|--|--------------|
| 1.0      | 15 Sep 17 | Initial Release  | L. Steven    |
| 2.0      | 06 Oct 19 | Procedure reviewed                                     | L. Steven    |
| 3.0      | 05 Jul 21 | Procedure reviewed and added to NTHSSA policy template | L. Steven    |
| 4.0      | 03 Jul 23 | Procedure reviewed                                     | L. Steven    |
|          |           |  |              |
|          |           |  |              |
|          |           |  |              |
|          |           |  |              |

DRAFT

**Disclaimer Message:** This is a **CONTROLLED** document for internal use only. Any documents appearing in paper form are not controlled and should be checked against the electronic file version prior to use.