

Stanton Territorial Hospital

FILENAME: COA10500 Sysmex CA660 Operator's Maintenance checklist.docx

P.O. Box 10, 550 Byrne Road YELLOWKNIFE NT X1A 2N1

Document Name: Sysmex CA660 Operator's Maintenance Checklist

Distribution: Core Laboratory

Document Number: COA10500

Date Issued: Feb 10, 2024

PRINT DATE: 3/13/2025 12:56:00 PM

Month/Year:						S	eria	l Nu	mbe	r:																					
DAILY MAINTENANCE	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Turn power																															
Wipe sample probe with alcohol swab																															
Turn ON																														1	
Check/replace distilled water																															
Check/empty waste container																														1	
Check/discard pneumatic trap chamber fluid																															
Empty/clean reaction tube trash box																															
Check & replenish reaction tubes																															
Load fresh OVB buffer as needed																															
Load fresh CA Clean 1 in position 1																															
Load fresh CA Clean 2 in position 2																															
Check/prepare reagents																															
Run "Rinse Probe"																															
Check temperatures																															
Cooler (13-17°C)																															
Detector (36-38°C)																															
Probe (36-38°C)																															
Room (15-35°C)																															
Remove dew from the reagent rack																															
Tech initials																															
WEEKLY MAINTENANCE	Date/Initial							Date/Initial						Date/Initial					Date/Initial						Date/Initial						
Clean instrument interior/exterior																															
Clean DI water rinse bottle with alcohol																															
								Co	mme	ents:																					
QUARTERLY MAINTENANCE				Date/	'Initia	l																									
Perform LED calibration																															
AS NEEDED MAINTENANCE				Date/	'Initia	I																									
Replace fuse																															
Clean filters																															
Check and drain trap chambers																															
Perform "Rinse and Prepare"									iina	rvico	r cia	nati	ıro.								Da	ato.									
								ر ا	upe	i vist	אוכ וי	iiail	ـ L. ت								0	ace							_		
YEARLY MAINTENANCE			Date/	'Initia	l																										
Replace rinse filter																															
NOTE: This is a CONTROLLED do version prior to use.	cume	ent fo	or int	ernal	use	only.	Any	docu	men	ts ap	pear	ing ir	pap	er fo	rm ar	e no	t con	trolle	d an	d sho	ould b	e ch	ecke	ed ag	ainst	elec	tronic	;			