



Stanton Territorial Hospital
P.O. Box 10, 550 Byrne Road
YELLOWKNIFE NT X1A 2N1

Document Name:
VITEK 2 Card Job Aid

Distribution:
Microbiology Instrumentation Manual

Document Number:
MIC70240

Date Issued:
3 June 2021

Uncontrolled When Printed

Card	McFarland Std.	Notes
GNI	0.48 to 0.63	
AST-N390	0.48 to 0.63	<ul style="list-style-type: none">• If mated GNI card not run, need to enter organism identification
GPI	0.48 to 0.63	
AST-GP67	0.48 to 0.63	<ul style="list-style-type: none">• If mated GPI card not run, need to enter organism identification
AST-ST03	0.48 to 0.63	<ul style="list-style-type: none">• If mated GPI card not run, need to enter organism identification
YST	1.80 to 2.20	
NH	2.70 to 3.30	
ANC	2.70 to 3.30	<ul style="list-style-type: none">• Need to enter offline tests: gram stain, morphology and aerotolerance

NOTE: This is a CONTROLLED document for internal use only. Any documents appearing in paper form are not controlled and should be checked against electronic version prior to use.

FILENAME: MIC70240-VITEK 2 Card Job Aid

PRINT DATE: 5/15/2025 2:08:00 PM