



TRAINING UPDATE

Lab Location: GEC
Department: Core

Date Distributed: 1/2/2013
Due Date: 1/31/2013

DESCRIPTION OF PROCEDURE REVISION

Name of procedure:

Salicylate by Dimension® Xpand Chemistry Analyzer GEC.C26 v002

Description of change(s):

Section	Reason
1, 7.1	Add analyzer name
5.3	Refer to PI for calibration levels
8, 16	Update SOP titles
8.2	Remove Lynx, specify Xpand process
10.4	Revise CRR lower limit
10.5	Revise Repeat Criteria, remove manual dilution
14.1	Revise AMR lower limit

Lower limit of CRR and AMR adjusted to match Vista analyzers (to maintain standardization)

Onboard dilution revised to match lower CRR, manual dilution eliminated

Other edits are minor

Document your compliance with this training update by taking the quiz in the MTS system.

Approved draft for training all sites (version 002)

Technical SOP

Title	Salicylate by Dimension® Xpand Chemistry Analyzer	
Prepared by	Ashkan Chini	Date: 4/12/2011
Owner	Robert SanLuis	Date: 11/9/2012

Laboratory Approval		Local Effective Date:
Print Name and Title	Signature	Date
<i>Refer to the electronic signature page for approval and approval dates.</i>		

Annual Review		
Print Name	Signature	Date

TABLE OF CONTENTS

1. Test Information.....3
 2. Analytical Principle4
 3. Specimen Requirements.....4
 4. Reagents.....5
 5. Calibrators/Standards5
 6. Quality Control7
 7. Equipment And Supplies9
 8. Procedure9
 9. Calculations.....11
 10. Reporting Results And Repeat Criteria.....11
 11. Expected Values.....12
 12. Clinical Significance.....12
 13. Procedure Notes12
 14. Limitations Of Method12
 15. Safety13
 16. Related Documents14
 17. References.....14
 18. Revision History14
 19. Addenda15

1. TEST INFORMATION

Assay	Method/Instrument	Local Code
Salicylate	Dimension® Xpand Chemistry Analyzer	SALIC

Synonyms/Abbreviations
ASA, Aspirin, SALIC

Department
Chemistry

Form revised 3/31/00

2. ANALYTICAL PRINCIPLE

The salicylate method is a modification of the Trinder colorimetric technique. Under acidic conditions, ferric nitrate forms a complex with salicylate. The amount of salicylate-ferric complex formed is proportional to the salicylate concentration and is measured using a two filter (510, 700 nm) endpoint with sample-reagent blanking technique.



3. SPECIMEN REQUIREMENTS

3.1 Patient Preparation

Component	Special Notations
Fasting/Special Diets	N/A
Specimen Collection and/or Timing	Normal procedures for collecting and storing serum may be used for samples to be analyzed by this method.
Special Collection Procedures	None
Other	N/A

3.2 Specimen Type & Handling

Criteria	
Type -Preferred -Other Acceptable	Serum None
Collection Container	Serum: Red top tube no additives
Volume - Optimum - Minimum	1.0 mL 0.5 mL
Transport Container and Temperature	Plastic vial or spun barrier tube transported at room temperature
Stability & Storage Requirements	Room Temperature: (20-25°C) 7 days
	Refrigerated: (2-8°C) 2 weeks
	Frozen: (-20°C or colder) 6 months
Timing Considerations	N/A
Unacceptable Specimens & Actions to Take	Specimens that are unlabeled, improperly labeled, or those that do not meet the stated criteria are unacceptable. Request a recollection and credit the test with the appropriate LIS English text code for "test not performed" message. Examples: Quantity not sufficient-QNS; Wrong collection-UNAC. Document the request for recollection in the LIS.
Compromising Physical Characteristics	Gross hemolysis. Reject sample and request a recollection. Credit the test with the appropriate LIS English text code.

Criteria	
Other Considerations	Allow to clot completely prior to centrifugation.

4. REAGENTS

Refer to the Material Safety Data Sheet (MSDS) supplied with the reagents for complete safety hazards. Refer to the section in this procedure covering “SAFETY” for additional information.

4.1 Reagent Summary

Reagents	Supplier & Catalog Number	Quantity
Salicylate	Siemens, Flex® reagent cartridge, Cat. No. DF20	4 Flex/carton

4.2 Reagent Preparation and Storage

NOTES: Date and initial all reagents upon opening. Each container must be labeled with (1) substance name, (2) lot number, (3) date of preparation, (4) expiration date, (5) initials of tech, (6) any special storage instructions; check for visible signs of degradation.

Refer to the Material Safety Data Sheet (MSDS) for a complete description of hazards. If a specific hazard is present, it will be noted in this procedure when the hazard is first encountered in a procedural step.

Irritant. Contains mixture of 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1).
Irritating to eyes and skin.
May cause sensitization by skin contact.
Avoid contact with skin.

Reagent	Salicylate
Container	Reagent cartridge
Storage	Store at 2-8° C
Stability	<ul style="list-style-type: none"> • Reagent is stable until expiration date stamped on the reagent cartridges. • Sealed or unhydrated cartridge wells on the instrument are stable for 30 days. • Once wells 1-8 have been entered by the instrument, they are stable for 5 days.
Preparation	Reagents are supplied ready for use. No additional preparation is required.

5. CALIBRATORS/STANDARDS

5.1 Calibrators/Standards Used

Calibrator	Supplier and Catalog Number
Salicylate Calibrator	Siemens Dimension®, Cat. No. DC38

5.2 Calibrator Preparation and Storage

NOTE: Date and initial all calibrators upon opening. Each container must be labeled with (1) substance name, (2) lot number, (3) date of preparation, (4) expiration date, (5) initials of tech (6) any special storage instructions; check for visible signs of degradation.

Calibrator	Salicylate Calibrator
Preparation	<ul style="list-style-type: none"> Allow to equilibrate at room temperature and mix thoroughly before use.
Storage/Stability	<ul style="list-style-type: none"> Store at 2-8° C Unopened calibrator is stable until expiration date stamped on the box. Open calibrator is stable for 3 months when vials are stored securely capped at 2-8°C.

5.3 Calibration Parameter

Criteria	Special Notations
Reference Material	Salicylate Calibrator
Assay Range	0.2 – 100 mg/dL
Suggested Calibration Level	See Reagent Package Insert for lot specific assigned values in mg/dL
Frequency	<ul style="list-style-type: none"> Every new reagent cartridge lot. Every 3 months for any one lot When major maintenance is performed on the analyzer. When control data indicates a significant shift in assay.
Calibration Scheme	Three levels in triplicate
Assigned Coefficients	C ₀ -1.330 C ₁ 0.666

5.4 Calibration Procedure

1. From Operating Menu press F5:Process Control press F1: Calibration Enter Password press F2: SETUP and RUN
2. Select the test method to be calibrated - if lot number is incorrect Press F1: Other Lot

3. Enter all information on screen
4. Press F8: QC yes/no to change to yes
5. Press F4: Assign cups If additional methods need to be calibrated, select the method.
6. Press F7: Load/run
7. Load cups into assigned position
8. Press F4: RUN

5.5 Tolerance Limits

IF.....	THEN.....
If result fall within assay-specific specification, and QC values are within acceptable limits,	proceed with analysis
If result falls outside assay-specific specification, or QC values are out of Acceptable limits,	troubleshoot the assay and/or instrument and repeat calibration

6. QUALITY CONTROL

6.1 Controls Used

Controls	Supplier and Catalog Number
Liquichek Immunoassay Plus Control Levels 1, 2 & 3	Bio-Rad Laboratories Cat # 361, 362 & 363

6.2 Control Preparation and Storage

NOTE: Date and initial all controls upon opening. Each container should be labeled with (1) substance name, (2) lot number, (3) date of preparation, (4) expiration date, (5) initials of tech, and (6) any special storage instructions; check for visible signs of degradation.

Control	Liquichek Immunoassay Plus Control Levels 1, 2 & 3
Preparation	Allow the frozen control to stand at room temperature (18-25°C) until completely thawed. Swirl the contents gently to ensure homogeneity. (Do not use a mechanical mixer) Use immediately. After each use, promptly replace the stopper and return to 2-8°C storage.
Storage/Stability	Open controls are stable for 14 days at 2-8°C. Thawed and unopened controls are stable for 30 days at 2-8°C. Unopened controls are stable until the expiration date at -20 to -70°C.

6.3 Frequency

Analyze all levels of QC material after every calibration and each day of testing.

Refer to the Dimension Xpand® QC Schedule in the Laboratory policy Quality Control Program and in the Dimension X-pand® Quick Reference Guide.

6.4 Tolerance Limits

Step	Action
1	Acceptable ranges for QC are programmed into the Laboratory Information System (LIS), and may be posted near the instrument for use during computer downtime.
2	Run Rejection Criteria <ul style="list-style-type: none"> Anytime the established parameters are exceeded (if one QC result exceeds 2 SD), the run is considered out of control (failed) and patient results must not be reported. The technologist must follow the procedure in the Laboratory QC Program to resolve the problem.
3	Corrective Action: <ul style="list-style-type: none"> All rejected runs must be effectively addressed through corrective action. Steps taken in response to QC failures must be documented. Patient samples in failed analytical runs must be <u>reanalyzed according to the Laboratory QC Program</u>. Supervisors may override rejection of partial or complete runs only with detailed documentation and criteria for overrides that are approved by the Medical Director. Consult corrective action guidelines in Laboratory QC Program. Follow corrective action guidelines in the Laboratory QC Program. Corrective action documentation must follow the Laboratory Quality Control Program.
4	Review of QC <ul style="list-style-type: none"> QC must be reviewed weekly by the Group Lead or designee and monthly by the Supervisor/Manager or designee. If the SD and/or CV are greater than established ranges, investigate the cause for the imprecision and document implementation of corrective actions.

6.5 Review Patient Data

Technologist must review each result print-out for error messages. Refer to the Dimension® system manual “Error messages” section for troubleshooting. Check for unusual patterns, trends, or distributions in patient results (such as an unusually high percentage of abnormal results). Resolve any problems noted before issuing patient reports.

6.6 Documentation

- QC tolerance limits are programmed into the instrument and the LIS. The LIS calculates cumulative mean, SD and CV and stores all information for easy retrieval.
- Quality control records are reviewed daily at the bench, weekly by the Group Lead or designee, and monthly by the Supervisor/Manager or designee.
- Refer to complete policies and procedures for QC documentation and for record retention requirements in the Laboratory QC Program.

6.7 Quality Assurance Program

- Each new lot number of reagent or new shipment of the same lot of reagent must be tested with external control materials and previously analyzed samples. Performance of the new lot must be equivalent to the previous lot; utilize published TEA for acceptability criteria.
- Training must be successfully completed and documented prior to performing this test. This procedure must be incorporated into the departmental competency assessment program.
- The laboratory participates in CAP proficiency testing. All proficiency testing materials must be treated in the same manner as patient samples.
- Monthly QC must be presented to the Medical Director for review and signature.
- Monthly QC mean and SD are sent to Bio-Rad Laboratories for peer group comparison.
- Consult the Laboratory QC Program for complete details.

7. EQUIPMENT and SUPPLIES

7.1 Assay Platform

Dimension [Xpand® System](#)

7.2 Equipment

- Refrigerator capable of sustaining 2–8°C.
- Freezer capable of sustaining range not to exceed -20 to -70°C.
- Centrifuge

7.3 Supplies

- Plastic serum tubes and serum cups
- Purified water (Millipore® or equivalent)
- Calibrated pipettes and disposable tips

8. PROCEDURE

SAL Flex® reagent cartridge Cat. No. DF20 is required to perform this test.

Salicylate is performed on the [Dimension Xpand® System](#) after the method is calibrated (see Reference Material in Calibration section) and Quality Controls are acceptable.

NOTE: For all procedures involving specimens, buttoned lab coats, gloves, and face protection are required minimum personal protective equipment. Report all accidents to your supervisor.

The package insert for a new lot of kits must be reviewed for any changes before the kit is used. A current Package Insert is included as a Related Document.

8.1	Instrument Set-Up Protocol
1.	For instrument set up and operation: Refer to Startup and Maintenance, Siemens Dimension® Xpand procedure .
2.	Check reagent inventory
3.	Sampling, reagent delivery, mixing, processing, and printing of results are automatically performed by the Dimension® Xpand system. For details of the automated parameters, see below under “Test conditions.”

8.2	Specimen/Reagent Preparation
1.	Centrifuge the specimens.
2.	Specimens are placed in Dimension® Xpand segments for analysis by the instrument. Refer to the Sample Processing, Siemens Dimension® Xpand procedure . The sample container (if not a primary tube) must contain sufficient quantity to accommodate the sample volume plus 50 µL of dead volume. Precise container filling is not required.

8.3	Specimen Testing
1.	For QC placement and frequency, refer to the Dimension® Xpand QC Schedule in the Laboratory QC Program.
2.	Follow the instructions, outlined in the Dimension® Xpand Operators Manual
3.	The instrument reporting system contains error messages to warn the user of specific malfunctions. Results followed by such error messages should be held for follow-up. Refer to the Dimension® Xpand system manual “Error messages” section for troubleshooting.
4.	Follow protocol in Section 10.5 “Repeat criteria and resulting” for samples with results above or below the Analytical Measurement Range (AMR). Repeat critical values and document according to Critical Values procedure. Investigate any failed delta result and repeat, if necessary.
5.	Append the appropriate English text code qualifier messages to any samples requiring a comment regarding sample quality and/or any other pertinent factors.

Test Conditions	
Sample Size:	15 µL
Reagent 1 Volume:	50 µL
Reagent 2 Volume:	50 µL

Form revised 3/31/00

Reagent 3 Volume:	50 µL
Test Temperature:	37° C
Wavelength:	510 and 700 nm
Type of measurement:	Bichromatic endpoint

9. CALCULATIONS

The instrument automatically calculates the concentration of Salicylate in mg/dL.

10. REPORTING RESULTS AND REPEAT CRITERIA

10.1 Interpretation of Data

None required

10.2 Rounding

No rounding is necessary. Instrument reports results to one decimal point.

10.3 Units of Measure

mg/dL

10.4 Clinically Reportable Range (CRR)

1.7 - 300.0 mg/dL

10.5 Repeat Criteria and Resulting

All repeats must replicate the original result within the total allowable error (TEa) of the assay. Refer to TEa listing for specific information.

Values that fall within the AMR or CRR may be reported without repeat. Values that fall outside these ranges must be repeated.

IF the result is ...	THEN...
< 1.7 mg/dL	Assure there is sufficient sample devoid of bubbles, cellular debris, and/or fibrin clots. Report as: < 1.7 mg/dL
≥ 100.0 mg/dL	On Board Automated Dilution: Results ≥ 100.0 mg/dL will automatically have repeat testing performed into the instrument using dilution factor of 3. No multiplication is necessary.
> 300.0 mg/dL	If the recommended dilution does not give results within the clinically reportable range, report as: "> 300.0 mg/dL-REP" Bring to the attention of your supervisor prior to releasing result.

Message	Code
Verified by repeat analysis	Append -REP to the result.

11. EXPECTED VALUES

11.1 Reference Ranges

2.8 – 20.0 mg/dL

11.2 Critical Values

> 30.0 mg/dL

11.3 Priority 3 Limit(s)

None established

12. CLINICAL SIGNIFICANCE

The SAL method used on the Dimension® Xpand clinical chemistry system is an *in vitro* diagnostic test intended to measure salicylates, a class of analgesic, antipyretic and anti-inflammatory drugs (including aspirin) in human serum. Salicylate test results may be used in the diagnosis and treatment of salicylate overdose and for monitoring salicylate levels during therapy.

13. PROCEDURE NOTES

- **FDA Status:** FDA Approved/cleared
- **Validated Test Modifications:** None

The instrument reporting system contains error messages to warn the operator of specific malfunctions. Any report slip containing such error messages should be held for follow-up. Refer to your Dimension Xpand Operator’s Guide.

A system malfunction may exist if the following 5-test precision is observed:

Concentration	S.D.
20 mg/dL	>1.0 mg/dL
100 mg/dL	>2.0 mg/dL

14. LIMITATIONS OF METHOD

14.1 Analytical Measurement Range (AMR)

1.7 – 100.0 mg/dL

14.2 Precision

Material	Mean mg/dL	Standard Deviation (%CV)	
		Within-run	Between-day
Serum Pool			

Level 1	25.9	0.22	0.26
Level 2	46.6	0.22	0.37
Level 3	88.6	0.33	0.76

14.3 Interfering Substances

Sodium azide at a concentration of 0.05% increases salicylate results by 20 mg/dL. Bilirubin (unconjugated) of 20 mg/dL at a SAL concentration of 33 mg/dL decreased SAL results by 20%. Lipemia (Intralipid®) of 600 mg/dL and above at a SAL concentration of 33 mg/dL tripped a test report message; therefore the magnitude of the interference could not be determined.

HIL Interference:

The SAL method was evaluated for interference from hemolysis, icterus and lipemia according to CLSI EP7-P. Bias, defined as the difference between the control sample (does not contain interferent) and the test sample (contains interferent), is shown in the table below. Bias exceeding 10% is considered “interference”.

Substance tested	Test Concentration SI Units	SAL concentration mg/dL	Bias %
Hemoglobin (hemolysate)	1000 mg/dL	32.9	<10
Bilirubin (unconjugated)	5 mg/dL	33.0	<10
Lipemia Intralipid®	200 mg/dL	32.6	<10

14.4 Clinical Sensitivity/Specificity/Predictive Values

Not available

15. SAFETY

The employee has direct responsibility to avoid injury and illness at work. Nearly all harmful exposures to infectious substances and chemicals, and other injuries, can be avoided with effective training and consistent safe work practices.

Become familiar with the Environmental Health and Safety (EHS) Manual to learn the requirements on working safely and protecting the environment from harm. Although lab work typically focuses on the hazards of working with specimens and chemicals, we must also control other important hazards.

- Slips, trips, and falls cause many serious injuries. Please ensure that spills are cleaned quickly (to avoid slippery floors) and that you can see and avoid obstacles in your path.
- Ergonomic injuries result from performing tasks with too much repetition, force, or awkward position. Ergonomic injuries include strains and back injuries. Learn about ergonomic hazards and how to prevent this type of injury.

- Scratches, lacerations, and needlesticks can result in serious health consequences. Attempt to find ways to eliminate your risk when working with sharp materials.

Report all accidents and injuries immediately to your supervisor or the business unit Environmental Health and Safety Manager or Specialist.

16. RELATED DOCUMENTS

1. Dimension Xpand® Clinical Chemistry System Operator's Manual
2. Calibration / Verification Siemens Dimension® Xpand procedure
3. Dimension Xpand® Cal Accept Guidelines
4. Dimension Xpand® Calibration summary
5. Sample Processing, Siemens Dimension® Xpand procedure
6. Start up and Maintenance, Siemens Dimension® Xpand procedure
7. Laboratory Quality Control Program
8. QC Schedule for Siemens Dimension Xpand®
9. Laboratory Safety Manual
10. Material Safety Data Sheets (MSDS)
11. Siemens Dimension Xpand® Limits Chart
12. Quest Diagnostics Records Management Procedure
13. Dimension Xpand® System Error Messages Chart
14. Centrifuge Use, Maintenance and Functions Checks (Lab policy)
15. Hemolysis, Icteria and Lipemia Interference (Lab policy)
16. Repeat Testing Requirements (Lab policy)
17. Critical Values (Lab policy)
18. Current Allowable Total Error Specifications at http://questnet1.qdx.com/Business_Groups/Medical/qc/docs/qc_bpt_tea.xls
19. Current package insert SAL Flex® Reagent Cartridge DF20

17. REFERENCES

1. Package Insert, SAL Flex® Reagent Cartridge DF20, Siemens Healthcare Diagnostics Inc., 04/03/2008.
2. Package Insert, Salicylate Calibrator DC38, Siemens Healthcare Diagnostics Inc., 05/2008.
3. Package Insert, Liquichek Immunoassay Plus Controls, Bio-Rad Laboratories, 04/2010.

18. REVISION HISTORY

Version	Date	Section	Reason	Reviser	Approval
			Supersedes SOP C077.001		
000	04/25/12	6.7	Add use of TEA for lot to lot runs	L Barrett	J Buss
000	04/25/12	10.4	CRR edited to correct range.	J Buss	Dr Cacciabeve
000	04/25/12	10.5	Remove code QNSR	L Barrett	J Buss
000	04/25/12	15	Update to standard wording	L Barrett	J Buss
001	11/9/12		Update owner	L Barrett	R SanLuis

001	11/9/12	1, 7.1	Add analyzer name	L Barrett	R SanLuis
001	11/9/12	5.3	Refer to PI for calibration levels	A Chini	R SanLuis
001	11/9/12	8, 16	Update SOP titles	L Barrett	R SanLuis
001	11/9/12	8.2	Remove Lynx, specify Xpand process	A Chini	R SanLuis
001	11/9/12	10.4	Revise CRR lower limit	A Chini	R SanLuis
001	11/9/12	10.5	Revise Repeat Criteria, remove manual dilution	A Chini	R SanLuis
001	11/9/12	14.1	Revise AMR lower limit	A Chini	R SanLuis

19. ADDENDA

None