

TRAINING UPDATE

Lab Location: GEC, SGMC & WAH
Department: Core Lab

Date Distributed: 9/19/2016
Due Date: 10/1/2016
Implementation: 10/1/2016

DESCRIPTION OF PROCEDURE REVISION

Name of procedure:
ESR Stat Plus Maintenance Log AG.F212.1
Description of change(s):
Added line for performing tech to each shift and instruction at top This revised FORM will be implemented on October 1, 2016

Document your compliance with this training update by taking the quiz in the MTS system.

ESR Stat Plus Maintenance Log

- Germantown Emergency Center
- Shady Grove Medical Center
- Washington Adventist Hospital

Month: _____

Year: _____

Instrument Serial Number: _____

Place a check mark to indicate task is complete and record tech code/initials

Day Shift	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Wipe interior with Alcohol prep																															
Run Both Levels of Control																															
Tech																															

Evening Shift	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Wipe interior with Alcohol prep																															
Run Both Levels of Control																															
Tech																															

Night Shift	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Wipe interior with Alcohol prep																															
Run Both Levels of Control																															
Tech																															

Weekly Maintenance*	Date:	Date:	Date:	Date:	Date:
	Tech:	Tech:	Tech:	Tech:	Tech:

* For Weekly Maintenance instructions refer to Addenda A in Erythrocyte Sedimentation Rate by ESR STAT PLUS SOP

CORRECTIVE ACTION

Date	Tech	Problem	Corrective Action

Weekly review:	Weekly review:	Weekly review:
Weekly review:	Weekly review:	Monthly review: