

## TRAINING UPDATE

**Lab Location:** GEC, SGMC & WAH  
**Department:** Core Lab

**Date Distributed:** 4/19/2018  
**Due Date:** 5/10/2018  
**Implementation:** 5/1/2018

### DESCRIPTION OF REVISION

<b>Name of procedure:</b>
<b>Room Temperature Log, Microbiology AG.F314.1</b>
<b>Description of change(s):</b>
<p>Change upper temperature to 28C (currently it is 30C). Why? This is necessary to meet requirements of new C diff test that will be implemented soon</p> <p><b>This revised form was implemented on May 1, 2018</b></p>

**Document your compliance with this training update by taking the quiz in the MTS system.**

## Room Temperature Log - Microbiology

Acceptable range: **15 - 28C**

Month: _____	Year: _____	Thermometer ID: _____
--------------	-------------	-----------------------

**Interp:** S = satisfactory, temp w/in acceptable range  
 U = unsatisfactory, circle result and document corrective action; notify supervisor/Incharge immediately.  
 Complete Variance form. Additional corrective action should be documented on the back of the sheet.  
 Check Reset box to document device is reset daily

Date	Minimum Temp.	Maximum Temp.	Interp	Reset	Tech Code	Corrective action	Weekly Review
1							
2							
3							
4							
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							

Monthly review: \_\_\_\_\_

Date: \_\_\_\_\_

**Corrective Action Codes:**

- |   |   |
|---|---|
| <p><b>A. Rechecked temperature</b> (<i>Log repeat temperature</i>)</p> <p><b>B. Adjusted Thermostat</b> (<i>Retake temperature in 1 hour</i>)</p> <p><b>C. Contacted Plant Ops / Facilities</b></p> | <p><b>D. Affected contents relocated</b></p> <p><b>E. Contents quarantined pending evaluation</b></p> <p><b>F. Other:</b> (<i>Document on reverse side</i>)</p> |
|---|---|