#### TRAINING UPDATE

Lab Location: Department:

GEC, SGMC & WOMC Core lab, Field Ops

Date Distributed:
Due Date:
Implementation:

1/28/2020 2/28/2020 **2/1/2020** 

#### **DESCRIPTION OF PROCEDURE REVISION**

## Name of procedure:

Refrigerator Temperature and Maintenance Form AG.F395.2

Freezer Temperature and Maintenance Form AG.F396.2

# **Description of change(s):**

These logs are being revised to standardize the format of documenting temperatures as cited during CQA inspection.

**Internal thermometer** must be read to nearest half degree (.0 or .5)

**Digital thermometer** must be read to one decimal place (.0 or .1 or .2 and so on)

These revised Logs will be implemented on February 1, 2020

Document your compliance with this training update by taking the quiz in the MTS system.



## **Refrigerator Temperature / Maintenance Form**

Shady Grove Medical Cente
White Oak Medical Cente
Germantown Emergency Cente

ID		Serial Number:			Interp: S = satisfactory, temp w/in acceptable range			
Mode <u>l</u>		Мо	Month/Year: U = un			isfactory, circle result and document corrective action;		
						notify super	rvisor immediately	
CHECK DAILY							Quarter	y Maintenance
Date	Tech Code	Internal Thermometer 2-8°C Read to nearest half degree (.0)	Digital Temp 2-8°C Read to one decimal place. Put N/A if Not	Interp S / U	Corrective Action  Use back of page if more space is needed	Weekly Review		arterly maintenance not due current month.  High Activation
		or .5)	Applicable				Temp. Alarm Test	Low Activation
1							romp. / warm root	□ Acceptable
2							(High < 7.5 and Low	□ Unacceptable
3							>2.5)	If unacceptable, document
4								resolution on back.
5								□ Acceptable □ N/A
6							Power Failure	□ Unacceptable
7							Alarm	If unacceptable, document resolution on back.
8								
9								□ Acceptable
							Clean Condenser Grill	·
11 12							Onn	If unacceptable, document resolution on back.
13								□ Acceptable
14							Wipe Clean and	□ Unacceptable
15							Inspect Door	If unacceptable, document
16							Gasket	resolution on back.
17								□ Acceptable □ N/A
18							Probe Bottle	□ Unacceptable
19							Inspection	If unacceptable, document
20								resolution on back.
21							Calibrate	□ Acceptable □ N/A
22							Temperature Probe	□ Unacceptable
23								The temperatures must agree within 1.0°C
24						ļ	Therm.	_
25 26							Digital	If unacceptable, document resolution on back.
27							Annual Maintenan	ce (check N/A if not due)
28								□ Acceptable □ N/A
29							Bleach Probe and	·
30							Deska Dattla	If unacceptable, document
31								resolution on back.
Mont	thly Revi	ew:				Date:		

### **Corrective Action Codes:**

- A. Rechecked temperature (Log repeat temperature)
- Adjusted Thermostat (Retake temperature in 1 hour) В.
- C. Contacted repair service

- D. Contents removed and equipment taken out of service
- E. Other: (Document on reverse side of this sheet)

AG.F395.2 Revised 1/2020



### Freezer Temperature / Maintenance Form

Shady Grove Medical Center
White Oak Medical Center
Germantown Emergency Center

F	re	ezer	Range	(check	one	١.
	16	CZCI	Nange	CHECK	OHE	,.

		Freezer Range (cneck one):
ID _	Serial Number:	□ -20 to -50C □ -20 to -25C GEC only
Mode	Month/Year:	□ -15 to -25C
Interp:	S = satisfactory, temp w/in acceptable range. U = ur	nsatisfactory, circle result & document corrective action; notify supervisor immediately

CHECK DAILY					Quarter	y Maintenance		
Date	Tech Code	Internal Thermometer Read to nearest	Digital Temp Read to one decimal place.	Interp	Corrective Action Use the back of the page if more space is	Weekly Review		arterly maintenance not due current month.
1 2 3 4	Code	half degree (.0 or .5)	Put N/A if Not Applicable	S/U	needed	Review	Temp. Alarm Test (check appropriate)  □ ≤ -20C or □ ≤ -15C	Activation Temp   Acceptable  Unacceptable  If unacceptable, document resolution on back.
5 6 7 8							Power Failure Alarm	□ Acceptable □ N/A □ Unacceptable  If unacceptable, document resolution on back.
9 10 11 12							Clean Condenser Grill and Inspect External Drain Fan	□ Acceptable □ Unacceptable  If unacceptable, document resolution on back.
13 14 15 16							Wipe Clean and Inspect Door Gasket	□ Acceptable □ Unacceptable If unacceptable, document resolution on back.
17 18 19 20							Probe Bottle Inspection	□ Acceptable □ N/A □ Unacceptable  If unacceptable, document resolution on back.
21 22 23 24 25							Calibrate Temperature Probe Therm Digital	□ Acceptable □ N/A □ Unacceptable  The temperatures must agree within 1.0°C  If unacceptable, document
26 27 28							Annual Maintenan	resolution on back.  ce (check N/A if not due)  □ Acceptable □ N/A
29 30 31							Bleach Probe and Probe Bottle	·

/lonthly Review:	Date:				
		,			

#### **Corrective Action Codes:**

- Rechecked temperature (Log repeat temperature)
- В. Adjusted Thermostat (Retake temperature in 1 hour)
- C. Contacted repair service

- D. Contents removed and equipment taken out of service
- **E.** Other: (Document on reverse side of this sheet)

AG.F396.2 Revised 1/2020