

TRAINING UPDATE

Lab Location:

GEC and FWMC

Date Distributed:

3/28/2022

Department:

Core Lab and FWMC Field
Ops

Due Date:

4/28/2022

DESCRIPTION OF PROCEDURE REVISION

Name of procedure:

Title: Millipore (AFS – Analyzer Feed System), Siemens Dimension® EXL (GEC.C279)

Description of change(s):

SOP:

- Updated instructions for how to order monthly culture for bacterial content of Millipore water (GEC and FWMC Core)
- Added instructions specific to FWMC on how to track for send out to Chantilly. (FWMC Field Ops and Core lab)

Form: Millipore AFS–16D Maintenance Log and System Readings- (AG.F539.4)

- Updated form to include all acceptable reference ranges.
- Added instructions: Values MUST be recorded in boxes on form.

The revised Form will be implemented **April 1, 2022**. The SOP is effective immediately.

Document your compliance with this training update by taking the quiz in the MTS system.

Non-Technical SOP

Title	Millipore (AFS – Analyzer Feed System), Siemens Dimension® EXL	
Prepared by	Valerie Rose	Date: 10/15/2020
Owner	Robert SanLuis	Date: 10/15/2020

Laboratory Approval		
Print Name and Title	Signature	Date
<i>Refer to the electronic signature page for approval and approval dates.</i>		
Local Issue Date:		Local Effective Date:

TABLE OF CONTENTS

1. PURPOSE.....	1
2. SCOPE.....	1
3. RESPONSIBILITY.....	1
4. DEFINITIONS.....	2
5. PROCEDURE.....	2
6. RELATED DOCUMENTS.....	10
7. REFERENCES.....	10
8. REVISION HISTORY.....	11
9. ADDENDA AND APPENDICES.....	11

1. PURPOSE

The purpose of this procedure is to outline the use and maintenance of the Millipore AFS Water Purification System. The Millipore AFS Water Purification System will be utilized as the Analyzer Feeder System for Dimension EXL analyzers.

2. SCOPE

This procedure applies to Laboratory staff who operate the Dimension EXL analyzers.

3. RESPONSIBILITY

Laboratory staff who operate the Dimension EXL analyzer, prepare reagents, utilize water or culture the system are responsible for monitoring and/or maintenance of the Millipore system.

4. DEFINITIONS

Clinical Laboratory Reagent Water (CLRW) – water that meets the requirements of the Clinical and Laboratory Standards Institute (CLSI).

AFS – Analyzer Feed System; a water purification system that produces CLSI Type I Product Water directly from potable tap water. The system combines Reverse Osmosis (RO) and Ion Exchange technology to accomplish this. The product water then feeds the analyzer directly, without any extra user interaction. The FAS system is designed to produce CLSI Type I quality water.

5. PROCEDURE

5.1 Operating principle and status

An Analyzer Feed System (“AFS®”) is a water system designed to automatically feed an analyzer. The AFS® D (or Degasser) includes a degassing device which is used to remove air from the product water delivered to the analyzer.

As a summary, the system purifies tap water using the Progard® Pack and Reverse Osmosis (RO). A tank and pump are used to store and distribute the product water to the Analyzer. A sensor is used to determine if the Analyzer needs water. This water is further purified by ion exchange before going to the Analyzer.

Understanding the terminology of operating status:

DISTRIBUTION -- The system has detected that the analyzer needs water. Thus the distribution part of the system is producing water and sends that water to the analyzer.

RECIRCULATION -- The system is periodically recirculating water in the distribution portion of the system but no water is being delivered to the analyzer.

TANK FILLING -- The distribution tank is being filled by RO Water. A sensor starts and stops this mode.

FLUSH -- A periodic flush of the RO portion of the system is done. This refreshes the water in this part of the system if they are not being used to produce water.

RINSING -- A sensor has determined that the RO portion needs to rinse itself before sending water to the tank.

SANITIZATION -- There are different types of sanitization that can be done. See the maintenance section in this document for more information.

CLEANING -- An acid or base cleaning can be done. A packaged chemical is available for this type of cleaning. See the maintenance section in this document for more information.

5.2 General Information

Parameter	CLSI/CAP specification for Type I Water
Maximum bacterial content	< 10 CFU/ml*
Minimum resistivity	10 MΩ-cm**
Maximum silicate level	< 0.05 mg/L
Particulate matter	0.22 micron filter
Organic contaminants	Activated carbon*** or Reverse Osmosis or Distillation
pH	Not specified

*CFU/ml = colony forming units per ml


** MΩ.cm=Megohm.cm compensated to 25° C

*** Not required by CAP

5.3 System Operation

5.3.1 Keypad Usage

The keypad allows you to interface with the system. Changing the software modes, reviewing the system's performance, and initiating maintenance procedures can be done via the keypad.



1 – MODE key – This button is used to switch between different system modes: Standby or maintenance mode, Ready or tank filling mode and Configuration mode.

- From Standby, press twice to go the Ready.
- From Ready, press once to go to Standby.


2 – LEFT or RIGHT key – These buttons are used to navigate in the user interface software. They are also used to change configuration or set point values.

3 – VALIDATE key – This button is used to confirm a change or an action required by the water system. It is also used to save any changes made in configuration.




4 – HELP INDICATOR icon – There are up to 3 of these icons on the display depending upon the software mode. These indicators are here to tell the user which keypad button is available and active.

5.3.2 Display Icons

Know how to interpret each icon displayed. These are displayed below.



1 – System status indications are:

- a.  Maintenance or tool: the system is in Standby or maintenance can be started.
- b.  Configuration or gears: configuration parameters such as set points can be changed.
- c.  Process or hourglass: the system is busy. No action can be done.

2 – **Warning maintenance or alarm** are displayed to indicate that maintenance action is required or that a parameter is out of range.


3 – **Values area** is used to display water quality measures, software menu and configuration values, or error code numbers.

























4 – **Units of measurement**

5 – **Consumables status** icons are used to indicate that a maintenance action is required.

6 – **Tank filling** is represented by the top bar. When the top bar is blinking, the system fills the tank. When it's not blinking, the tank full level is reached.

5.3.3 View Operation Parameters

In Ready or Tank filling, press and hold 2 seconds on left and right arrows  to access operation parameters.


	Tap water feed conductivity value can be displayed by pressing  Press  to go back to d1 view.	
	RO feed water conductivity value can be displayed by pressing  Press  to go back to d2 view.	
	RO water temperature value can be displayed by pressing  Press  to go back to d3 view.	
	RO pump pressure value can be displayed by pressing  Press  to go back to d4 view.	
	RO % rejection value can be displayed by pressing  Press  to go back to d5 view.	
	RO permeate water conductivity value can be displayed by pressing  Press  to go back to d6 view.	


5.3.4 Sanitization and Cleaning

A chlorine tablet is periodically used to reduce or remove a layer of bio film on the feed water side of the RO membrane. It may also become necessary to remove any build up layer of mineral scale or organic material, which is done with a pH cleaner.

Refer to section 5.3.7 of this document, or to the Maintenance section of the User Manual, for a complete description of both cycles.

5.3.5 Error Messages and Basic Troubleshooting

An error code is always displayed with the  icon and can be displayed in two different ways:







Yellow display alert error: 

Red display alarm error: 




Please see the User Manual for the list of error codes and basic troubleshooting.


5.3.6 Replacement of Consumables

Table for quick reference for when the consumables should be exchanged:




Consumables names	When to change the Consumables
Progard pack	 <p>When the system show this icon, order a new Progard pack. It will be exhausted in 15 days.</p>  <p>When the system displays this yellow alert, replace the Progard pack. It is recommended to replace the Tank Vent filter at this time. It is also recommended to clean the feed water strainer. Additionally, replace the Progard if the system has been turned off for more than 48 hours. System should be powered on and in STANDBY mode when the Progard is replaced.</p>
Tank Vent filter	Replace the Tank Vent filter when the Progard is exchanged.
Q-Gard pack	 <p>When the system displays this icon, order a new Q-Gard. It will be exhausted in 15 days.</p>  <p>When the system displays this yellow alert, replace the Q-Gard.</p>
BioPak C final filter	 <p>When the system displays this icon, order a new final filter. It will be exhausted in 15 days.</p>  <p>When the system displays this yellow alert, replace the final filter.</p>



a) Replacing Progard Pack

- When the system shows the icon  order a new pack.
- When the system displays the yellow alert  replace the pack (it is recommended to replace the Tank Vent filter at this time, and to clean the feed water strainer).
- Verify or put the system in Standby by pressing the Mode key 
- To REMOVE the exhausted pack:






1. Unlock the locking handle by pulling up.
 2. Pull out the pack.
Note: when the pack is removed, a RED color display appears. This is normal. This will go away when a new pack is installed.
- To place a NEW pack:
 1. Remove the new pack from its shipping box. Remove the 2 protective caps from the ports at the top of the pack. Locate the O-ring inside each port to make sure the O-rings are pushed inside against the inner edge of the ports.
 2. Mount the bottom of the pack onto the small hole in the cabinet. Swing the rest of the pack so that the top of the Pack aligns with pack locking handle and the cabinet ports.
 3. Push the top of the pack in place.
 4. Push the pack locking handle down to secure the pack.
 - The new pack is detected.
 - The display prompts you to start a “Progard flush”: press  for 1 second.
 - The “Progard flush” lasts 15 minutes.
 - System goes in Tank Filling automatically.

b) Replacing Q-Gard Pack

- When the system shows the icon  order a new pack.
- When the system displays the yellow alert  replace the pack.
- BEFORE replacing the pack, verify the tank is FULL.
- Verify or put the system in Standby by pressing the Mode key .
- To REMOVE the exhausted pack:
 1. Locate the Sampling Port. It is typically located downstream of the final filter. Place a container under to collect any water that will come out when it's opened.
 2. Open the valve. Wait until water stops flowing out.
 3. Detach the tubing from the final filter inlet. Place this tubing end into a drain or sink.
 4. Pull up the locking handle.
 5. Pull out the entire pack.
Note: when the pack is removed, an alarm appears. This is normal. The alarm will go away when a new pack is installed.
- To place a NEW pack:
 1. Remove the new pack from its shipping box. Remove the 2 protective caps from the ports at the top of the pack. Locate the O-ring inside each port to make sure the O-rings are pushed inside against the inner edge of the ports.
 2. Mount the bottom of the pack onto the small hole in the cabinet. Swing the rest of the pack so that the top of the Pack aligns with pack locking handle and the cabinet ports.
 3. Push the top of the pack in place.
 4. Push the pack locking handle down to secure the pack.



- The new pack is detected.
 - The display prompts you to start a “polisher rinse”: press  next for 1 second.
 - Wait for 2-3 minutes.
 - Put the system in Standby.
5. Reconnect the tubing on the final filter inlet.
 6. Close the sampling valve. Press  to put the system in Ready or Tank Filling.



c) Replacing Final Filter


- When the system shows the icon  order a new final filter.
- When the system displays the yellow alert  replace the final filter.
- Before replacing the final filter verify that the tank is at least HALF FULL.
Note: do not rinse the final filter with a pack that has not been rinsed out.
- To REMOVE the final filter:
 1. Verify or put the system in Standby by pressing .
 - The system displays LF: 01.
 2. Open the sampling valve to remove the pressure in the distribution product line.
 3. Disconnect the final filter from the distribution product line.
 4. Unscrew the fittings from both ends of the filter.
 5. Remove the final filter.
- To place a NEW final filter:
 1. Remove the new filter from its shipping box.
 2. Screw the fittings on both ends.
 3. Connect the new filter on the distribution product line, and open or verify that the sampling valve is opened.
 4. Press  for 1 second to put the system in Tank Filling and to reset the life time.
 - Remove air from the final filter by unscrewing the air vent cap.
 - Close it once it is air purged.
 - The display prompts you to start a final filter rinse: press  next for 1 second.
 - Wait for 2-3 minutes.
 - Put the system in Standby.
 5. Close the sampling valve.

5.3.7 Maintenance






a) Sanitizing RO Membranes with Chlorine

- A chlorine tablet is periodically used to reduce or remove a layer of bio film on the feed water side of the RO membrane.
- When the system shows the yellow alert , put the system in Standby by pressing 

1. Using the tool, unscrew the sanitation port cap (refer to the Quick Reference Guide for visual aid).
2. Remove the plug and insert one chlorine tablet into the port using proper safety equipment.
3. Screw the plug back into place, making sure the O-ring is in place.
4. To start the CL2 sanitization, press  for 1 second to start CL2 sanitization. You will see this icon 
 - The sanitization lasts 19 minutes, then the system goes into Ready or Tank Filling.

Note: The sanitization can be cancelled in emergency cases or when started by mistake. Press  during the cycle to abort sanitization.

b) Cleaning RO Membranes with acid/base (pH)

- A pH cleaning should be performed as needed, as it may become necessary to remove any build up layer of mineral scale or organic material. This is done with a pH Cleaner.
- To start the cleaning, make sure the system is in Standby or put the system in Standby by pressing 
 1. Using the tool, unscrew the sanitation port cap (please refer to the Quick Reference Guide for visual aid).
 2. Remove the plug and insert one chlorine tablet into the port.
 3. Screw the plug back into place, making sure the O-ring is in place.
 4. To start the pH cleaning, with the system already in Standby, press on RIGHT arrow  to go past the CL2 icon to this icon 
 5. Press  for 1 second to start pH cleaning.
 - The cleaning lasts 82 minutes.
 6. Once the cleaning is finished, press  to put the system in Standby.
 7. Unscrew the plug and REMOVE the empty pouch from the cleaning port.
 8. Screw the plug back and put the system in Tank Filling by pressing twice.

5.3.8 Monitoring Bacterial Levels in Water

To meet CLSI requirements for the specification of bacteria count, periodic testing is recommended. Bacteria cultures must be performed monthly.

a) To obtain water for culture:

- 1) Sanitize the sampling port (upper port with RED cap): remove red cap and wash the port with a 10% bleach solution and then with an alcohol prep pad.
- 2) Open the sampling port SLOWLY to avoid being sprayed and let it run for at least 1 minute.
- 3) Collect 10 ml of water in a sterile container (urine culture collection cup).
- 4) Record collection of water (date/tech code) on the AFS-16(D) Daily Maintenance Log

b) To order and process the culture:

- 1) Order via the LIS:

- a. Log in to Sunquest through **Order Entry**
 - b. For Look up Mode: Patient ID
 - c. Hospital number: type **MILL-1** (Millipore, Monthly)
 - d. Enter **Collection date, Collection time, Receive date, and Receive time**
 - e. Order Physician: press ENTER to take default 99987
 - f. Order Code: type **XH2O** and for
 - g. SDES: type **WATER**.
 - h. Label the containers using the LIS accession labels.
- 2) Prepare for culture at SGMC:
- a. In the LIS, track the order to SGMC using template code GIC.
 - b. **Sample must be REFRIGERATED within 2 hours of collection** (stable for 24 hours).
 - c. Send to SGMC for FES and subsequent transport to Chantilly for culture.
- 3) Prepare for culture at FWMC:
- a. In LIS, ROB (interface 602) the order.
 - b. **Sample must be REFRIGERATED within 2 hours of collection** (stable for 24 hours).
 - c. Send to WOMC for FES and subsequent transport to Chantilly for culture.
- c) Culture results:
- 1) Access and print results from LIS one week after sending.
 - 2) Record results as CFU/mL on the AFS-16(D) Daily Maintenance Log
 - 3) Acceptable values are ≤ 10 CFU/mL
 - 4) Document corrective action if value is unacceptable:
 - a. Collect another sample and re-submit for testing
 - b. If repeat test is unacceptable, contact Millipore

5.3.9 Documentation of Maintenance

Complete AFS-16(D) Maintenance Log and System Readings form

6. RELATED DOCUMENTS

AFS-16(D) Maintenance Log and System Readings (AG.F539)

7. REFERENCES

User manual, Millipore, V 4.0, 04/2019. www.millipore.com,
www.millipore.com/techservice, www.millipore.com/lab_water
1-888-645-5478

CLSI Document GP40-A4-AMD, Vol.26, No.22 (formerly C03-A4-AMD), Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Standards 4th Edition, 2012

8. REVISION HISTORY

Version	Date	Reason for Revision	Revised By	Approved By
1	3/15/21	Section 5.3.7: changed hospital number for SQ ordering	L Barrett	R SanLuis
2	7/29/21	Header: added FWMC	L Barrett	R SanLuis
3	3/23/22	Section 5.3 fixed numbering	D Collier	R SanLuis
3	3/23/22	Section 5.3.8 revised culture ordering process and added FWMC instructions to prepare for tracking.	M Sabonis	R SanLuis

9. ADDENDA

None

For Training only Not for printing or binding