

ACL TOP - ANALYTICAL MEASUREMENT RANGE (AMR) CHART

- St. Joseph Medical Center Tacoma, WA
 St. Clare Hospital Lakewood, WA
 St. Elizabeth Hospital Enumclaw, WA
 St. Francis Hospital Federal Way, WA
 St. Anthony Hospital Gig Harbor, WA
 PSC

Test	Analytical Measurement Range (AMR)	Acceptable Diluent	Maximum Manual Dilution	Lower Limit report as	Upper Limit report as
Activated Partial Thromboplastin Time (PTT)	16 – 270 SEC	N/A	N/A	<16	>270 SEC
D-Dimer HS (DD HS)	150 - 69,000 ng/mL	Factor Diluent	N/A	<150	>69,000 ng/mL
Fibrinogen (FIB)	35 - 1000 mg/dL	N/A	N/A	<35	>1000 mg/dL
Low Molecular Weight Heparin (LMWH)	0.04 – 2.00 IU/mL	N/A	N/A	<0.04	>2.00 IU/mL
Prothrombin Time (PT)	8 - 320 SEC	N/A	N/A	<8	>320 SEC
Unfractionated Heparin (UFH)	0.04 – 2.00 IU/mL	N/A	N/A	<0.04	>2.00 IU/mL
Thrombin Time (TT)	5 - 300 SEC	N/A	N/A	<5	>300 SEC

DOCUMENT APPROVAL Purpose of Document / Reason for Change:		
1/14/14: Updated upper limit of aPTT to 270.		
<input checked="" type="checkbox"/> <i>No significant change to process in above revision. Per CAP, this revision does not require further Medical Director approval.</i>		
Committee Approval Date	<input type="checkbox"/> Date: 7/2/15 <input type="checkbox"/> N/A – revision of department-specific document which is used at only one facility	Medical Director Approval <i>(Electronic Signature)</i> <i>Katie Wilkinson, MD</i> 7/30/15