

IRIS VELOCITY QC & MAINTENANCE LOG- SA, SC, SF

- | | | |
|--|---|--|
| <input type="checkbox"/> St. Joseph Medical Center Tacoma, WA | <input checked="" type="checkbox"/> St. Clare Hospital Lakewood, WA | <input type="checkbox"/> St. Elizabeth Hospital Enumclaw, WA |
| <input checked="" type="checkbox"/> St. Francis Hospital Federal Way, WA | <input checked="" type="checkbox"/> St. Anthony Hospital Gig Harbor, WA | <input type="checkbox"/> Highline Medical Center Burien, WA |
| | | <input type="checkbox"/> PSC |

INSTRUMENT: Iris iChem™ Velocity Urine Chemistry Analyzer **MONTH:** _____ **YEAR:** _____

QUALITY CONTROL

Reagent Strip Lot#	Manuf Exp. Date:	Date in Use:	Tech
CA/CB/CC Lot#	Manuf Exp. Date:	Date in Use:	Tech

DAILY MAINTENANCE

Date	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Clean waste container																															
Check strip supply																															
Check wash bottle																															
Clean belts, surfaces																															
CA, CB, & CC QC																															
TECH																															
When Strips added- clean SPM, new dessicant																															

WEEKLY MAINTENANCE	Date:				Comments: Reviewed by: Date:
Clean CGM					
TECH					
MONTHLY MAINTENANCE	Date:				
Clean the strip conveyer system (SCS)					
Clean wash station bath					
TECH					
QUARTERLY MAINT (Feb, May, Aug, Nov)	Date:				
Reflectance CalChek					
SG/CC CalCheck					
TECH					
AS NEEDED MAINTENANCE	Date:				
Clean Tube Detector, Barcode reader, & optical sensors					
TECH					