

 <p>STOP Notification</p>	<h2 style="text-align: center;">Upper Limit Analytic Measurement Range (AMR) change for Prealbumin from 60 mg/dL to 55 mg/dL</h2>	
<p>When:</p>	<p style="text-align: center;">Tuesday, June 23, 2020</p>	
<p>What?</p>	<p>Due to inability of Moses Cone Hospital to verify AMR of 60 mg/dL, the upper limit AMR range has been reduced from 60 mg/dL to 55 mg/dL.</p>	
<p>Who is Affected:</p>	<p style="text-align: center;">Affected Location(s)</p> <ul style="list-style-type: none"> <input checked="" type="checkbox"/> Alamance Cancer Center <input checked="" type="checkbox"/> High Point Cancer Center <input checked="" type="checkbox"/> Wesley Long Cancer Center <input checked="" type="checkbox"/> Alamance Regional <input checked="" type="checkbox"/> Annie Penn Hospital <input checked="" type="checkbox"/> MedCenter @ High Point <input checked="" type="checkbox"/> MedCenter @ Mebane <input checked="" type="checkbox"/> Moses Cone Hospital <input checked="" type="checkbox"/> Wesley Long Hospital Women's Hospital 	<p style="text-align: center;">Department(s)</p> <ul style="list-style-type: none"> Blood Bank Cytology Flow Cytometry Histology Microbiology Phlebotomy Point Of Care <input checked="" type="checkbox"/> Rapid Response Lab Respiratory Therapy Specimen Processing
<p>Why?</p>	<p>Moses Cone was not able to verify current upper limit AMR for prealbumin.</p>	
<p>What you will need to do to prepare:</p>	<p>Sites performing prealbumin testing will have to adjust their respective analyzers with the new AMR.</p>	
<p>Manager / Supervisor Responsibility:</p>	<p style="text-align: center;">Send Post Live Documentation to IT Manager within 5 days of the effective date for test systems marked with an "X"</p>	
	<p style="text-align: center;">Sunquest</p>	<p style="text-align: center;">WindowPath</p>
	<p style="text-align: center;">CHL</p>	<p style="text-align: center;">Instrument / Manual Test</p>
	<p style="text-align: center;">PowerPath</p>	
<p style="text-align: center;">Each sites supervisor/manager or designee will be responsible for ensuring analyzer(s) prealbumin AMR is updated on effective date.</p>		
<p>Need Help?</p>	<p style="text-align: center;">Contact your Manager / Supervisor</p>	