

Procedure Change/Revision (IMM 36-A)

Procedure: **Beta2 Glycoprotein IgA on the DSX** (Doc#IMM 123)

Change/Revision: To the Quality control section the following will be added:

Assess the quality of the standard curve and review the predicted concentrations of the standards. The R Squared value for the standard curve must be >0.97.

Date of Revision:

1/11/13

Director Signature/Date:

Handwritten signature and date: 1/11/2013

Technologist Signature/Date:

See MTS