

Procedure Change/Revision (IMM 36-A)

Procedure: PSA Access (Doc# IMM 75)

Change/Revision:

Analytical Measurement Range has been corrected from 0.1-153.0 ng/mL to the manufactures suggested range of 0.008 -153.0 ng/mL.

The reportable lower limit of detection has been changed from 0.1 ng/mL to 0.008 ng/mL. This reflects the true sensitivity of the Hybritech PSA assay and eliminates the need to perform a separate ultra sensitive PSA assay. All values less than 0.008 will be reported as <0.008 ng/ mL.

Date of Revision: 4/9/12

Director Signature/Date:

Handwritten signature and date: *B. [unclear]*
4/10/12

Technologist Signature/Date: