

# Procedure Change/Revision (IMM 36-A)

**Procedure: Indirect Platelet Antibody(IPLT)**

**Change/Revision:**

**Positive Control should have an OD of  $\geq 1.0$ , previously was  $\geq 1.4$ .**

**Negative control should be  $\leq 0.16$  for all three glycoproteins: GPIIb/IIIa, GPIb/IX, and GPIa/IIa.**

**Date of Revision: 6/6/2013**

**Director Signature/Date:**

Handwritten signature and date: M. J. W. 6/7/13

---

**Technologist Signature/Date:**