

Principle

Intended Use

CREAm reagent, when used in conjunction with UniCel® DxC 800 System and SYNCHRON® Systems AQUA CAL 1 and 2, is intended for the quantitative determination of creatinine concentration in human serum, plasma, urine or fluids.

Clinical Significance

Creatinine measurements are used in the diagnosis and treatment of renal diseases, in monitoring renal dialysis, and as a calculation basis for measuring other urine analytes.

Methodology

The UniCel® DxC System(s) determine creatinine concentration by means of the Jaffe rate method.(1)

A precise volume of sample (16.5 microliters serum or 5.5 microliters urine) is injected in a reaction cup containing an alkaline picrate solution. The ratio used is one part sample to 35 parts reagent for serum and one part sample to 105 parts reagent for urine. Creatinine from the sample combines with the reagent to produce a red color complex. Absorbance readings are taken at 520 nanometers between 19 and 25 seconds after sample injection. The absorbance rate has been shown to be a direct measure of the concentration of creatinine in the sample.(2,3,4)

Chemical Reaction Scheme

Creatinine + Picric Acid \longrightarrow Creatinine-Picrate Complex (red)

Specimen

Acceptable Sample Containers

13 x 75 PST, SST and Red Top BD tubes

PST, SST and Red Top BD microtainers

Spot urines should be aliquoted into a 13 x 75 clear cap BD tube

24 hour urine collections are usually received in 3000 ml plastic urine collection jugs.

Fluids should be received in a 13 x 75 Clear Cap BD tube

Type of Specimen

Biological fluid samples should be collected in the same manner routinely used for any laboratory test.(5)

Freshly drawn serum, plasma, fluids or properly collected urine (random/timed) are the preferred specimens.

Whole blood is not recommended for use as a sample.

Specimen Storage and Stability

Tubes of blood are to be kept closed at all times and in a vertical position. It is recommended that the serum or plasma be physically separated from contact with cells within two hours from the time of collection.(5)

Separated serum or plasma should not remain at room temperature longer than 8 hours. If assays are not completed within 8 hours, serum or plasma should be stored at +2°C to +8°C. If assays are not completed within 48 hours, or the separated sample is to be stored beyond 48 hours, samples should be frozen at -15°C to -20°C. Frozen samples should be thawed only once. Analyte deterioration may occur in samples that are repeatedly frozen and thawed.(5)

It is recommended that random urine assays be performed within 2 hours of collection. (6,7)



For timed specimens, the collection container is to be kept in the refrigerator or on ice during the collection period. No preservative is required. Upon receipt in the laboratory, the collection container must be stored refrigerated until testing is performed. **Testing must be performed within 4 days (96 hours) of start of collection.**(6,7,8,9)

Sample Volume

The optimum volume, when using a 0.5 mL sample cup, is 0.3 mL of sample. For optimum primary sample tube volumes and minimum volumes, refer to the [Primary Tube Sample Template](#).

Unacceptable Specimens

Urine samples aspirated by the IRIS IQ are not acceptable for urine chemistry testing.

Urine collected in a BD UA Preservative Tube is not acceptable for urine chemistry testing.

Refer to the [Procedural Notes](#) section of this chemistry information sheet for information on unacceptable specimens.

Reagents

Contents

Each kit contains the following items: **Kit Reorder # 472525**

- Two Alkaline Buffer Bottles (1600 mL)
- Two Picric Acid Solution Bottles (400 mL)
- Instruction Insert

Volumes per Test

Sample Volume	Serum 16.5 µL Urine 5.5 µL
Total Reagent Volume	570 µL

Reactive Ingredients

Reagent Constituents

Alkaline Buffer:

Sodium Hydroxide 0.188 mol/L

Picric Acid Solution:

Picric Acid 0.05 mol/L

Also non-reactive chemicals necessary for optimal system performance.

CAUTION!

Avoid skin contact with reagent. Use water to wash reagent from skin.

Materials Needed But Not Supplied With Reagent Kit

- SYNCHRON® Systems AQUA CAL 1 and 2
- At least two levels of control material
- Saline

Reagent Preparation

1. Carefully pour 400 mL of Picric Acid Solution into the 1600 mL Alkaline Buffer bottle. Replace cap and mix at least 10 times by gentle inversion.
2. Record preparation date on the bottle.
3. If excessive foam is produced when mixing, allow foam to dissipate before loading.
4. Freshly prepared creatinine reagent may contain micro air bubbles that may result in calibration failure or calibration with low span. To prevent this phenomenon, allow the prepared reagent to sit with the cap loosened for a minimum of 30 minutes (or overnight) before loading onto the instrument.
5. Date and initial reagent container and document in reagent log before loading each new carboy.

NOTICE

Do not reuse old reagent or mix fresh reagent with old reagent.

Acceptable Reagent Performance

The acceptability of this reagent is determined by successful calibration and by ensuring that quality control results are within acceptance criteria, as defined in the Clinical Chemistry Quality Control Procedure #3000.T.

Reagent Storage and Stability

Alkaline Buffer and Picric Acid Solution stored unopened and unmixed at room temperature are stable until the expiration dates indicated on each bottle. The combined Creatinine Reagent is stable on-instrument for 30 days from the date of preparation. **Do not freeze or refrigerate.** If reagent is frozen in transit, thaw completely, warm to room temperature and mix thoroughly by gently inverting bottle a least 10 times.

NOTICE

At reduced temperature, a precipitate may form in the Alkaline Buffer or combined Creatinine Reagent. Do not filter the precipitate. DO NOT USE combined Creatinine Reagent until all precipitate is completely redissolved. It will redissolve upon warming to 21°C to 25°C without any loss of reactivity. A 25°C water bath may be used to warm reagent. Mix after redissolving precipitate by inverting bottle 10 times.

Equipment

This test is performed on the Beckman UniCel DxC 800 systems; Beckman-Coulter, Brea, California. For technical assistance, call the Beckman-Coulter hotline: 1-800-854-3633.

Refer to the UniCel DxC 800 systems [Reference Manual](#) for detailed instructions.

Calibration

Calibrator Required

SYNCHRON LX AQUA CAL 1 and 2. ([Kit Reorder #s 471288 and 471291](#))

Calibrator Preparation

No preparation is required.

Calibrator Storage and Stability

If unopened, the calibrators should be stored at 2°C to 8°C until the expiration date printed on the calibrator bottle. Once opened, the calibrators are stable at room temperature for 30 days.

Repetitive refrigeration of the aqueous calibrators may facilitate crystal formation. Once removed from refrigerated storage, these calibrators should remain at room temperature.

Calibration Information

The system must have a valid calibration in memory before controls or patient samples can be run.

Under typical operating conditions the creatinine assay must be calibrated every 72 hours or with each new bottle of reagent and also with certain parts replacements or maintenance procedures, as defined in the UniCel DxC 800 Systems [Instructions for Use](#) (IFU) manual.

For calibration instructions, refer to the UniCel DxC 800 System [Instructions For Use](#) (IFU) manual.

The system will automatically perform checks on the calibration and produce data at the end of calibration. In the event of a failed calibration, the data will print out with error codes and the system will alert the operator of the failure. The explanation of these error codes can be found in the UniCel DxC 800 System [Instructions For Use](#) (IFU) manual.

Traceability

Creatinine (analyte) in this calibrator is traceable to IDMS*.

* IDMS - Isotope Dilution Mass Spectroscopy.

The set point values were established based upon the gravimetric addition of specific quantities of the measurand to achieve the appropriate concentration.

The values are verified using representative samples from each lot of calibrator and are specific to the assay methodologies of the SYNCHRON systems. Values determined by other methodologies may be different. Such differences, if present, may be caused by inter-method bias.

Quality Control

At least two levels of control material should be analyzed each shift. In addition, these controls should be run with each new calibration, with each new reagent cartridge, and after specific maintenance or troubleshooting procedures as detailed in the UniCel DxC 800 System [Instructions For Use](#) manual. More frequent use of controls or the use of additional controls is left to the discretion of the user based on workload and workflow.

The following controls should be used in accordance with the package instructions for use inserts. Quality control results should be evaluated and handled with respect to the Clinical Chemistry Quality Control Procedure #3000.T. Controls are compiled statistically in the LIS and reagent lot changes are documented on DxC Reagent Log sheets.

Quality Control Material

Control	Storage
MAS ChemTrak 1	+2°C to +8°C*
MAS ChemTrak 3	+2°C to +8°C*
MAS Urine Chemistry Control 1	+2°C to +8°C**
MAS Urine Chemistry Control 2	+2°C to +8°C**

*Controls are received frozen and stored at -15°C to -25°C. Bottles of controls in use are thawed and stored at +2°C to +8°C and are good for 14 days.

**Bottles of controls are stored at +2°C to +8°C. Once opened, controls are good for 30 days when stored at +2°C to +8°C.

Testing Procedure

1. If necessary prepare reagent as defined in the Reagent Preparation section of this chemistry information sheet and load the reagent onto the system.
2. After reagent load is completed, calibration is required.
3. Program samples and controls for analysis.
4. After loading samples and controls onto the system, follow the protocols for system operation.

For detailed testing procedures, refer to the UniCel DxC 600/800 System [Instructions For Use](#) (IFU) manual

Calculations

UniCel DxC Systems perform all calculations internally to produce the final reported result. The system will calculate the final result for sample dilutions made by the operator when the dilution factor is entered into the system during sample programming.

If the dilution was programmed in Remisol, the final calculated result from a dilution will not be calculated by the UniCel DxC system but by Remisol.

Timed Urine Specimens

Calculated from the following equations:

24 hr timed urine specimens

$$\text{Urine Creatinine} \frac{\text{mg}}{\text{dL}} \times \frac{\text{dL}}{100\text{mL}} \times \text{Total volume collected (mL)} = \text{mg/24hr}$$

Calculations are only performed on 24 hour collections (± 15 minutes) and reported as mg/24hr. Do not round off total collection time.

For creatinine clearance, the following equation is used:

$$\frac{\text{Urine Creatinine}}{\text{Serum Creatinine}} \times \frac{\text{Total volume collected (mL)}}{\text{Total time of collection (mins)}} = \text{uncorrected} \times \frac{\text{Std BSA}}{\text{BSA of pt}} = \text{corrected creatinine clearance} \quad \text{mL/min}$$

Standard Body Surface Area (BSA) = 1.73

BSA (A) calculation: $\log A = (0.425 \log W (\text{weight})) + (0.0725 \log H (\text{height})) - 2.144$

Use total collection time in minutes. Do not round off total collection time.

Estimated Glomerular Filtration Rate (eGFR)

Calculated for certain patient populations using the IDMS-Traceable Modification of Diet in Renal Disease (MDRD) Study equation:

$\text{GFR (mL/min/1.73 m}^2) = 175 \times (\text{S}_{\text{cr}})^{-1.154} \times (\text{Age})^{-0.203} \times (0.742 \text{ if female}) \times (1.210 \text{ if African American})$
(conventional units)

The National Kidney Disease Education Program (NKDEP) suggests the use of an estimating or prediction equation to estimate glomerular filtration rate from serum creatinine for people with chronic kidney disease (CKD) and those at risk for CKD (diabetes, hypertension, cardiovascular disease, and family history of kidney disease). A suggested equation for this purpose is the MDRD Study equation. This equation requires four variables: IDMS Traceable serum/plasma creatinine, age in years (18 years or older), gender, and race (validated for Caucasian and African- American populations). Primary reasons for these recommendations are:

- GFR and creatinine clearance are poorly inferred from serum creatinine alone. This is mainly because these are related inversely (non-linearly) to serum creatinine. The effects of age and gender, and to a lesser extent race, on creatinine production further inhibit effective interpretation.
- Creatinine is more often measured than urinary albumin. For patients with diabetic nephropathy, increased urinary albumin excretion often occurs before decreases in GFR. However, serum creatinine is measured frequently and may be the initial screening test for CKD.
- Measurement of kidney function (GFR or creatinine clearance) is essential once albuminuria is present.
- The MDRD equation is the most thoroughly validated equation. Further validation is under way in additional populations (people with normal GFR, people with diabetes, and Hispanics).
- The MDRD equation is superior to other methods of approximating GFR. Direct comparison of the MDRD equation to other equations such as Cockcroft-Gault and even 24-hour urine collections has proven this superiority.
- The normal serum creatinine reference interval does not necessarily reflect a normal GFR for an individual patient. Primary care providers and other specialists should routinely use an estimating equation to assess patients' kidney function.
- The MDRD equation does not require weight or height variables. The equation yields a GFR result normalized to 1.73m² body surface area, which is an accepted average adult body surface area.

eGFR Reference Interval:

> 60 mL/min./1.73 square meters

The NKDEP recommends reporting eGFR values **greater than or equal to 60 mL/min/1.73 m²** simply as "≥60 mL/min/1.73 m²," not as an exact number.

For values **below 60 mL/min/1.73 m²**, the report should give the numerical estimate rounded to a whole number (e.g., "32 mL/min/1.73 m²").

The equation has been most extensively evaluated in people with chronic kidney disease and reduced GFR and is less accurate for persons with normal or mildly impaired kidney function.

Quantification of eGFR values of 60 mL/min/1.73 m² and below have more clinical implications for classification of kidney function than values above this level.

Note: The estimated GFR result assumes a steady-state and is most accurate for GFRs <60mL/min./1.73m². The eGFR is not reliable in certain groups, including severely ill patients. Also, patients > 59 years of age can have a mildly reduced GFR due to aging. The MDRD equations used to estimate GFR have been validated only in Caucasians and African-Americans 18 - 70 years of age. The equations have not been validated in other population groups, including pregnant women, transplant recipients, medically unstable patients including those with acute renal failure, or in persons with extremes of body size, muscle mass, or nutritional status. Application of the MDRD calculation in these cases may lead to errors in GFR estimation.

For more information, refer to the NKDEP website: <http://www.nkdep.nih.gov>

Reporting Results

Equivalency between the SYNCHRON LX and UniCel DxC 800 Systems has been established. Chemistry results between these systems are in agreement and data from representative systems may be shown.

Reference Intervals

The following reference intervals were taken from literature and a study performed on SYNCHRON Systems.⁽¹⁰⁾

Reference Intervals

Intervals	Sample Type	Conventional Units	S.I Units
Literature	Serum or Plasma (Male)	0.9 - 1.3 mg/dL	80 - 115 µmol/L
	Serum or Plasma (female)	0.6 - 1.1 mg/dL	53 - 97 µmol/L
	Urine (Male)	800 - 2000 mg/24hrs	7.1 - 17.7 mmol/24hrs
	Urine (Female)	600 - 1800 mg/24hrs	5.3 - 15.9 mmol/24hrs
SYNCHRON	Serum or Plasma (Male)	0.64 - 1.27 mg/dL	57 - 113 µmol/L
	Serum or Plasma (female)	0.44 - 1.03 mg/dL	39 - 91 µmol/L
UCDMC	Serum or Plasma (Male)	0.64 - 1.27 mg/dL	57 - 113 µmol/L
	Serum or Plasma (female)	0.44 - 1.03 mg/dL	39 - 91 µmol/L
	Urine - 24 hr	1000 - 2000 mg/24hrs	8.8 - 17.7 mmol/24hrs
	Urine - Creatinine Clearance	70 - 130 ml/min	1.17 - 2.17 ml/s

Refer to References (11,12,13) for guidelines on establishing laboratory-specific reference intervals.

Reference interval for a spot or random urine sample has not been established.

A reference interval for fluids has not been determined by UCDHS. This result may be less accurate than for the usual sample type, and should be interpreted in the context of the patient's clinical condition.

Procedural Notes

Anticoagulant Test Results

If plasma is the sample of choice, the following anticoagulants were found to be compatible with this method based on a study of 20 healthy volunteers:

Compatible Anticoagulants

Anticoagulant	Level Tested for In Vitro Interference	Average Plasma-Serum Bias (mg/dL) ^a
Ammonium Heparin	14 Units/mL	NSI
Lithium Heparin	14 Units/mL	NSI
Sodium Heparin	14 Units/mL	NSI
Potassium Oxalate/ Sodium Fluoride	2.0 / 2.5 mg/mL	NSI

^a NSI = No Significant Interference (within ± 0.2 mg/dL or 6%).

Limitations

If urine samples are cloudy or turbid, it is recommended that they be centrifuged before transfer to a sample cup.

Interferences

The following substances were tested for interference with this methodology:

Interferences

Substance	Source	Level Tested	Observed Effect ^a
Acetoacetic Acid	Acetoacetic Lithium Salt	5 mg/dL	+ 0.04 mg/dL ^b
		50 mg/dL	+ 0.4 mg/dL ^b
		125 mg/dL	+ 0.9 mg/dL
		500 mg/dL	+ 3.5 mg/dL
Bilirubin (unconjugated)	Bovine	20 mg/dL	- 0.2 mg/dL
Cefactor	NA ^c	100 ug/dL	+ 0.2 mg/dL
Cefoxitin	Cefoxitin sodium salt	50 µg/dL	+ 0.2 mg/dL
Cephalothin	NA	NA	+0.2 mg/dL
α-D-Glucose	NA	NA	+ 0.2 mg/dL
Fluorescein	Fluorescein Disodium Salt	220 mg/dL	Results suppressed
Glutathione	NA	1.5 mmol/L	+ 0.2 mg/dL
Hemoglobin	RBC hemolysate	500 mg/dL	NSI ^d
L-Dopa	NA	160 mg/dL	- 0.2 mg/dL
Lipemia	Intralipid ^e	500 mg/dL	NSI
	Human	Serum Index 8	NSI
Methyl dopa	NA	10 mg/dL	- 0.2 mg/dL
Pyruvic acid	NA	5 mg/dL	+ 0.2 mg/dL
Sulfasalazine	NA	60 mg/dL	NSI
Sulfobromophthalein	Sulfobromophthalein sodium salt	2.0 mg/dL	NSI

^a Plus (+) or minus (-) signs in this column signify positive or negative interference.

^b The observed effect at 5 and 50 mg/dL levels of acetoacetic acid are calculated based on the extrapolation of the interference data collected with 0, 125, 250, 375, and 500 mg/dL of acetoacetic acid.

^c NA = Not applicable.

^d NSI = No Significant Interference (within ±0.2 mg/dL or 6%).

^e Intralipid is a registered trademark of KabiVitrum, Inc., Clayton, NC 27250.

Grossly lipemic samples should be ultracentrifuged and the analysis performed on the infranate.

Refer to References (14,15,16,17) for other interferences caused by drugs, disease and preanalytical variables

Performance Characteristics

Analytical Measurement Range (AMR)

The SYNCHRON[®] System(s) method for the determination of this analyte provides the following analytical ranges:

Analytical Measurement Range

Sample Type	Conventional Units	S.I. Units
Serum/Plasma/Fluids	0.10 - 25.00 mg/dL	8.84 - 2210 µmol/L
Urine	10 - 400 mg/dL	0.88 - 35.36 mmol/L

Clinical Reportable Range (CRR)

(as determined on site):

Clinical Reportable Range

Sample Type	Conventional Units	S.I. Units
Serum/Plasma/Fluids	0.10 - diluted result mg/dL	8.84 - diluted result µmol/L
Urine	10 - diluted result mg/dL	0.88 - diluted result mmol/L

Serum/plasma/fluid samples with concentrations below the AMR and CRR (< 0.10 mg/dL) are reported as "**< 0.10 mg/dL**". Serum/plasma/fluid samples with concentrations greater than the AMR (> 25.0 mg/dL) are diluted with saline and reanalyzed.

Urine samples with concentrations below the AMR and CRR (< 10.0 mg/dL) are reported as "**< 10.0 mg/dL**". Urine samples with concentrations greater than the AMR (> 400.0 mg/dL) are diluted with saline and reanalyzed.

If the dilution was programmed in Remisol, the final calculated result from a dilution will not be calculated by the UniCel DxC system but by Remisol.

Sensitivity

Sensitivity is defined as the lowest measurable concentration which can be distinguished from zero with 95% confidence. Sensitivity for this analyte determination is 0.1 mg/dL (8.84 µmol/L) for serum or plasma and 10 mg/dL (0.88 mmol/L) for urine.

Equivalency

Equivalency was assessed by Deming regression analysis of patient samples to accepted clinical methods.

As determined by Beckman

Serum or Plasma (in the range of 7 to 130 mg/dL):

Y (SYNCHRON LX Systems)	= 0.964X - 0.02
N	= 97
MEAN (SYNCHRON LX Systems)	= 6.78
MEAN (SYNCHRON CX7 DELTA)	= 7.04
CORRELATION COEFFICIENT (r)	= 0.9993

Urine (in the range of 18.2 to 399 mg/dL):

Y (SYNCHRON LX Systems)	= 0.997X - 1.70
N	= 94
MEAN (SYNCHRON LX Systems)	= 135.8
MEAN (SYNCHRON CX7 DELTA)	= 138.0
CORRELATION COEFFICIENT (r)	= 0.9983

Serum or Plasma (in the range of 1.0 to 24.3 mg/dL):

Y (UniCel DxC Systems)	= 1.037X - 0.01
N	= 137
MEAN (UniCel DxC Systems)	= 2.8
MEAN (SYNCHRON LX Systems)	= 2.7
CORRELATION COEFFICIENT (r)	= 0.999

University of California, Davis Health System
Department of Pathology and Laboratory Medicine
Automated Chemistry/Urinalysis

Creatinine (CREAm) – Serum, Plasma, Urine, Fluids
Beckman UniCel DxC Systems

Technical Procedure 3122

Urine (in the range of 17.9 to 412.7 mg/dL):

Y (UniCel DxC Systems)	= 1.000X + 0.97
N	= 110
MEAN (UniCel DxC Systems)	= 136.1
MEAN (SYNCHRON LX Systems)	= 135.2
CORRELATION COEFFICIENT (r)	= 1.000

Serum (in the range of 4.42 to 22.45 mg/dL):

Y (UniCel DxC Systems)	= 1.01X - 0.03
N	= 39
MEAN (UniCel DxC Systems)	= 4.42
MEAN (IDMS* reference procedure (18))	= 4.40
CORRELATION COEFFICIENT (r)	= 0.9996

*IDMS = Isotope Dilution Mass Spectroscopy

Refer to References (19) for guidelines on performing equivalency testing.

Equivalency assessed by Deming regression analysis of patient samples to accepted clinical methods.

As determined at UCDMC

Serum or Plasma (in the range of 0.30 to 24.38 mg/dL):

Y (UniCel Dx800-4118)	= 0.962X + 0.031
N	= 58
MEAN (UniCel Dx800-4118)	= 2.475
MEAN (UniCel Dx800-1805)	= 2.541
CORRELATION COEFFICIENT (r)	= 0.9994

Urine (in the range of 27.30 to 372.40 mg/dL):

Y (UniCel Dx800-4118)	= 0.991X + 1.054
N	= 24
MEAN (UniCel Dx800-4118)	= 126.462
MEAN (SYNCHRON LX20PRO-2194)	= 126.579
CORRELATION COEFFICIENT (r)	= 0.9997

Serum or Plasma (in the range of 0.30 to 24.38 mg/dL):

Y (UniCel Dx800-4427)	= 0.966X - 0.004
N	= 58
MEAN (UniCel Dx800-4427)	= 2.450
MEAN (UniCel Dx800-1805)	= 2.541
CORRELATION COEFFICIENT (r)	= 0.9994

Urine (in the range of 27.30 to 372.40 mg/dL):

Y (UniCel Dx800-4427)	= 0.987X + 1.148
N	= 24
MEAN (UniCel Dx800-4427)	= 126.078
MEAN (SYNCHRON LX20PRO-2194)	= 126.579
CORRELATION COEFFICIENT (r)	= 0.9999

Serum or Plasma (in the range of 0.30 to 24.38 mg/dL):

Y (UniCel Dx800-4449)	= 0.969X + 0.041
N	= 57
MEAN (UniCel Dx800-4449)	= 2.536
MEAN (UniCel Dx800-1805)	= 2.575
CORRELATION COEFFICIENT (r)	= 0.9997

University of California, Davis Health System
Department of Pathology and Laboratory Medicine
Automated Chemistry/Urinalysis

Creatinine (CREAm) – Serum, Plasma, Urine, Fluids
Beckman UniCel DxC Systems

Technical Procedure 3122

Urine (in the range of 27.30 to 372.40 mg/dL):

Y (UniCel Dx800-4449)	= 1.007X + 0.499
N	= 24
MEAN (UniCel Dx800-4449)	= 128.013
MEAN (SYNCHRON LX20PRO-2194)	= 126.579
CORRELATION COEFFICIENT (r)	= 0.9999

Serum or Plasma (in the range of 0.24 to 23.30 mg/dL):

Y (UniCel Dx800-4427)	= 1.003X - 0.032
N	= 63
MEAN (UniCel Dx800-4427)	= 2.310
MEAN (UniCel Dx800-4118)	= 2.334
CORRELATION COEFFICIENT (r)	= 0.9999

Urine (in the range of 26.92 to 370.87 mg/dL):

Y (UniCel Dx800-4427)	= 0.996X + 0.098
N	= 24
MEAN (UniCel Dx800-4427)	= 126.078
MEAN (UniCel Dx800-4118)	= 126.462
CORRELATION COEFFICIENT (r)	= 0.9999

Serum or Plasma (in the range of 0.24 to 23.30 mg/dL):

Y (UniCel Dx800-4449)	= 1.007X + 0.012
N	= 62
MEAN (UniCel Dx800-4449)	= 2.391
MEAN (UniCel Dx800-4118)	= 2.362
CORRELATION COEFFICIENT (r)	= 0.9997

Serum or Plasma (in the range of 0.15 to 23.34 mg/dL):

Y (UniCel Dx800-4449)	= 1.004X + 0.044
N	= 62
MEAN (UniCel Dx800-4449)	= 2.391
MEAN (UniCel Dx800-4427)	= 2.338
CORRELATION COEFFICIENT (r)	= 0.9997

Urine (in the range of 26.92 to 370.87 mg/dL):

Y (UniCel Dx800-4449)	= 1.017X - 0.573
N	= 24
MEAN (UniCel Dx800-4449)	= 128.013
MEAN (UniCel Dx800-4118)	= 126.462
CORRELATION COEFFICIENT (r)	= 0.9997

Urine (in the range of 26.66 to 367.17 mg/dL):

Y (UniCel Dx800-4449)	= 1.021X - 0.672
N	= 24
MEAN (UniCel Dx800-4449)	= 128.013
MEAN (UniCel Dx800-4427)	= 126.078
CORRELATION COEFFICIENT (r)	= 0.9997

University of California, Davis Health System
Department of Pathology and Laboratory Medicine
Automated Chemistry/Urinalysis

Creatinine (CREAm) – Serum, Plasma, Urine, Fluids
Beckman UniCel DxC Systems

Technical Procedure 3122

Precision

A properly operating SYNCHRON® or UniCel DxC System(s) should exhibit imprecision values less than or equal to the maximum performance limits in the table below. Maximum performance limits were derived by an examination of the imprecision of various methods, proficiency test summaries, and literature sources.

As determined by Beckman

Maximum Performance Limits

Type of Precision	Sample Type	1 SD		Changeover Value ^a		%CV
		mg/dL	µmol/L	mg/dL	µmol/L	
Within-run	Serum/Plasma	0.1	9	3.3	300	3.0
Total	Serum/Plasma	0.2	13	3.3	300	4.5
Within-run	Urine	2.0	177	66.7	5900	3.0
Total	Urine	3.0	265	66.7	5900	4.5

^a When the mean of the test precision data is less than or equal to the changeover value, compare the test SD to the SD guideline given above to determine the acceptability of the precision testing. When the mean of the test precision data is greater than the changeover value, compare the test % CV to the guideline given above to determine acceptability. Changeover value = (SD guideline/CV guideline) x 100.

Precision established at UCDCM

Type of Precision	Sample Type	n	Mean (mg/dL)	1 SD	%CV
DxC800-4118 Within-run	SYNCHRON 1	20	0.994	0.029	3.0
	SYNCHRON 3	20	7.731	0.035	0.5
	MAS Urine Chemistry 1	20	71.137	0.599	0.8
	MAS Urine Chemistry 2	20	170.903	1.227	0.7
DxC800-4427 Within-run	SYNCHRON 1	20	0.968	0.028	2.9
	SYNCHRON 3	20	7.623	0.055	0.7
	MAS Urine Chemistry 1	20	70.575	0.420	0.6
	MAS Urine Chemistry 2	20	169.738	0.865	0.5
DxC800-4449 Within-run	SYNCHRON 1	20	1.034	0.023	2.2
	SYNCHRON 3	20	7.713	0.058	0.7
	MAS Urine Chemistry 1	20	70.385	0.471	0.7
	MAS Urine Chemistry 2	20	169.577	1.035	0.6

University of California, Davis Health System
Department of Pathology and Laboratory Medicine
Automated Chemistry/Urinalysis

Creatinine (CREAm) – Serum, Plasma, Urine, Fluids
Beckman UniCel DxC Systems

Technical Procedure 3122

Type of Imprecision	Sample Type	n	Mean (mg/dL)	SD	%CV
DxC800-4118 Day to Day	MAS ChemTrak 1	1323	0.75	0.024	3.2
	MAS ChemTrak 3	1327	6.55	0.093	1.4
	MAS Urine Chemistry 1	274	70.85	1.358	1.9
	MAS Urine Chemistry 2	269	167.43	3.203	1.9
DxC800-4427 Day to Day	MAS ChemTrak 1	1321	0.75	0.019	2.5
	MAS ChemTrak 3	1321	6.55	0.073	1.1
	MAS Urine Chemistry 1	372	70.68	1.212	17
	MAS Urine Chemistry 2	377	166.63	2.931	1.8
DxC800-4449 Day to Day	MAS ChemTrak 1	1336	0.75	0.048	6.4
	MAS ChemTrak 3	1333	6.57	0.091	1.4
	MAS Urine Chemistry 1	359	69.90	1.242	1.8
	MAS Urine Chemistry 2	382	165.67	3.037	1.8

Comparative Performance

The data for a SYNCHRON LX® System evaluated using the NCCLS Proposed Guideline EP5-T2 appears in the table below. (20) Each laboratory should characterize their own instrument performance for comparison purposes.

As determined by Beckman

NCCLS EP5-T2 Precision Estimate Method

Type of Imprecision	Sample Type		No. Systems	No. Data Points ^a	Test Mean Value (mg/dL)	EP5-T2 Calculated Point Estimates	
						SD	%CV
Within-run	Serum	Control 1	1	80	0.57	0.03	4.8
	Serum	Control 2	1	80	7.86	0.08	1.0
	Urine	Control 1	1	80	90.90	0.71	0.8
	Urine	Control 2	1	80	244.73	1.72	0.7
Total	Serum	Control 1	1	80	0.57	0.05	8.2
	Serum	Control 2	1	80	7.86	0.25	3.1
	Urine	Control 1	1	80	90.90	2.28	2.5
	Urine	Control 2	1	80	244.73	6.94	2.8

NOTICE

These degrees of precision and equivalency were obtained in typical testing procedures on a SYNCHRON LX® System and are not intended to represent the performance specifications for this reagent.

Additional Information

For more detailed information on UniCel DxC Systems, refer to the [Instructions for Use](#) and [Reference](#) manual.

References

1. Jaffe, M. Z., *Physiol. Chem.*, 10:391 (1886).
2. Bartels, E., Cikes, M., *Clin. Chem. Acta*, 26:1 10 (1969).
3. Fabiny, D. L., Ertingshausen, G., *Clin. Chem.*, 17:8 696 (1971).
4. Heinegard, D., Tiderstrom, G., *Clin. Chem. Acta*, 43:305 310 (1973).
5. Tietz, N. W., "Specimen Collection and Processing; Sources of Biological Variation", *Textbook of Clinical Chemistry*, 5th Edition, W. B. Saunders, Philadelphia, PA (2005).
6. National Committee for Clinical Laboratory Standards, *Procedures for the Handling and Processing of Blood Specimens Approved Guideline*, NCCLS publication H18-A, Villanova, PA (1990).
7. National Committee for Clinical Laboratory Standards, *Routine Urinalysis and Collection, Transportation and Preservation of Urine Specimens Tentative Guideline*, NCCLS publication GP16-T, Villanova, PA (1992).
8. Spierto, F.W., Hannon, W.H., Gunter, E.W., Smith, S.J., "Stability of Urine Creatinine", *Clin. Chim Acta*, 264(2):227-32 (1997).
9. van Berkel, E.A.T., Schel, O., Boer, A.K., "Influence of the Storage Temperature on Urine Analysis in Timed Samples", *Ned Tijdschr Klin Chem Labgeneesk*, 35:173-174 (2010).
10. Pagana, K D and Pagana, T J, *Mosby's Manual of Diagnostic and Laboratory Tests 3rd Edition*, Mosby Inc., St Louis, MO (2006).
11. National Committee for Clinical Laboratory Standards, *How to Define, Determine, and Utilize Reference Intervals in the Clinical Laboratory Approved Guideline*, NCCLS publication C28-A, Villanova, PA (1995).
12. Tietz, N. W., ed., *Fundamentals of Clinical Chemistry*, 6th Edition, W. B. Saunders, Philadelphia, PA (2007).
13. Henry, J. B., *Clinical Diagnosis and Management by Laboratory Methods*, 22nd Edition, W. B. Saunders Company, Philadelphia, PA (2006).
14. Young, D. S., *Effects of Drugs on Clinical Laboratory Tests*, 5th Edition, AACC Press, Washington, D. C. (2000).
15. Friedman, R. B., Young, D. S., *Effects of Disease on Clinical Laboratory Tests*, 4th Edition, AACC Press, Washington, D.C. (2001).
16. Young, D. S., *Effects of Preanalytical Variables on Clinical Laboratory Tests*, 3rd Edition, AACC Press, Washington, D. C. (2007).
17. De Leacy E. A., Brown N. N., Clague A. E., *Clin Chem* 35:1772-1774 (1989).
18. Stockl D, Reinauer H., *Clin Chem* 39:993-1000 (1993).
19. National Committee for Clinical Laboratory Standards, *Method Comparison and Bias Estimation Using Patient Samples, Approved Guideline*, NCCLS publication EP9-A, Villanova, PA (1995).
20. National Committee for Clinical Laboratory Standards, *Precision Performance of Clinical Chemistry Devices, Tentative Guideline*, 2nd Edition, NCCLS publication EP5-T2, Villanova, PA (1992).
21. CREm, SYNCHRON® System(s) Chemistry Information Sheet A18483AJ, (4/2013), Instruction for use sheet, https://www.beckmancoulter.com/wsrportal/techdocs?docname=/cis/A18483/AJ/EN_CREm.pdf

University of California, Davis Health System
 Department of Pathology and Laboratory Medicine
 Automated Chemistry/Urinalysis

Creatinine (CREAm) – Serum, Plasma, Urine, Fluids
 Beckman UniCel DxC Systems

Technical Procedure 3122

Prepared By	Date Adopted	Supersedes Procedure #
Michael Inn	October, 2000	Reformatted

Revision Date	Type of Revision	Revised by	Review/Annual Review Date	Reviewed By
			11/27/2000	G. Kost
			12/28/2001	G.Kost
			10/16/2002	G. Kost
			10/10/2003	S. Devaraj
			10/25/2004	S. Devaraj
			11/28/2005	G. Kost
			09/26/2006	G. Kost
			11/05/2007	G. Kost
			06/16/2008	G. Kost
July, 2008	IDMS traceability & eGFR	M. Inn		
			09/15/2009	G. Kost
			10/12/2010	G. Kost
12/2010	update	M.Inn		
06/28/2011	Added Fluids as a sample type	M.Inn	07/06/2011	G. Kost
			11/16/2011	G. Kost
08/07/2013	No reference interval for spot/random urines	M. Inn	08/16/2013	G. Kost
			09/17/2013	G. Kost
04/03/2015	Updated 24-hour collection stability	kdagang	04/15/2015	J. Gregg