



DXI Dilution Chart

CHEM.DXI.1.2 DXI Dilution Chart

TEST	QLS CODE	MISYS CODE	AMR	CRR	Pre-assigned On-board dil	Report Low as:	Report High as:	Dilute To endpoint?	Manual Dilution with:	units
AFPTM	237	AFPTM	0.5-3000	0.5-300,000	100	<0.5	>300,000	NO*		ng/mL
DHEA	402	DHEAS	2-1000	2-10,000	10	<2	>10,000	NO*		ug/dL
PTH	8837	PTH	1-3500	1-35000	10	<1	>35,000	NO*		pg/mL
CA125 (OV MONITOR)	29256	CA125	1-5000	1-100,000	20	<1	>100,000	NO*		u/mL
SHBG	30740	SHBG	0.5-200	0.5-2000	10	<0.5	>2000	NO*		nmol/L
THYROGLOBULIN	30278	THYR	0.1-500	0.1-5000	Manual Dilution X10	<0.1	>5,000	NO*	Thyroglobulin Sample Diluent	ng/mL
ANTI-TG AB (ATG)	267	THYAB	1-2500	N/A	NONE	<1	>2500	NO*		IU/mL
ANTI-TPO (ATA)	5081	TPABY	1-1000	1-10,000	10	<1	>10,000	NO*		IU/mL
<ul style="list-style-type: none"> • If a physician requests dilution to endpoint, Use higher on board dilution if available, otherwise perform manual dilution, and tag result with "PRDIL" (Physician requested dilution). • 										