



MONTH / YEAR: _____

Scale Calibration – A&D Digital Scale

In order to comply with industry standards, scale should be checked for optimal performance and cleaned at least four (4) times a year – quarterly. Additional checks should be done upon receipt of new equipment and after repair.

Scale Calibration

Date: _____

Weight	Expected Weight ($\pm 0.5\%$)	Scale Readings	Interpretation (S/U)
100g	99.5g – 100.5g		
200g	199.0g – 201.0g		
200g	199.0 – 201.0g		
500g	497.5 – 502.5g		
1000g	995.0g – 1005.0g		

Satisfactory Un-satisfactory, corrective action: _____

