



# i-STAT EC8+ AND CG8+ CARTRIDGES THEY LOOK SIMILAR BUT ARE NOT THE SAME



VS



What tests do they run?

	Na	K	Cl	CREA	Anion Gap calculated	Ionized Calcium	GLU	BUN	Hct	Hgb calculated	pH	pCO2	pO2	TCO2 calculated	HCO3 calculated	BE calculated	sO2 calculated
EC8+	X	X	X		X		X	X	X	X	X	X		X	X	X	
CG8+	X	X				X	X		X	X	X	X	X	X	X	X	X

What is the room temperature expiration?



**EC8+**  
**14 DAY**

Room  
Temperature  
Expiration  
Date

**CG8+**  
**2 MONTH**

Room  
Temperature  
Expiration  
Date

