

UniCel DxC 800
CALIBRATION REPORT
INSTRUMENT S/N: 003068

Calibrator: LDLD CAL Lot: M503300 Rack(s): 515

Rgt: LDLD mg/dL S/N: 66U Lot: 506154 Loc: 18
Cuv SetPts Blank Reaction Recov Run
57 0.0 0.0022 0.0019 -0.22 07/28/15
56 0.0 0.0022 0.0027 0.22 08:32:35
59 116.0 0.0055 0.1955 115.17
58 116.0 0.0063 0.1991 116.83
Cal Factor: 606.39 Offset: -0.07

REAGENT LOT S/N CONDITION

LDLD 506154 66U Calibrated