



Test information


Test code: 10TT-53	Test type: ExtendedRE	Sample ID: 10110074505	Rack ID: 05	Sample position: 1
Ordered date/time: 03/11/2020 04:25:57	Completed date/time: 03/11/2020 04:19:15	Status: FAILED	Upload status: Unloaded	Validation status: Calibrated

Results

Measured	Mean	Replicate 1	Replicate 2
Unit 1	FAILED s	FAILED s	
Unit 2	FAILED s	FAILED s	
Unit 3			
Unit 4			

Replicate 1 | Replicate 2 | HIL | Tracking Information

Display derivative curves



Errors and warnings:

Group	Code	Description
		Replicate 1
		Measured (s)
CE	5072	(Coag) Second derivative peak not found
		Unit 1 (s)
RE	5100	Measured result failed

LIS status:  Busy  Connected  Ready

Auto Run status:  Busy  Connected  Ready

