

Levey-Jennings Chart Period: 1/07/2016 - 28/07/2016

Lab: 421872, C1

Lot: 16780, Unassayed Chemistry

Alkaline Phosphatase

Cum Mean/SD/CV: [1] 80.7/2.4/2.95, [2] 339.5/7.6/2.24, **Fixed Mean/SD/CV: [1] 81.5/2.9/3.50, [2] 343.0/12.0/3.50**



Summary Stats for 1/07/2016 to 28/07/2016 Mean/SD/CV/Points: [1] 80.2/2.3/2.90/82, [2] 338.6/8.2/2.43/82

Graph against: Your laboratory, Current instrument, Evaluation Mean/SD

~~QC Sample~~

~~QC Repeat with fresh QC~~

~~QC Sample repeat run to assess analytical imprecision~~



Post Calibration

Red Flag

QC RUN WITH FRESH QC