

URMC Laboratories and Affiliates at Strong West
156 West Ave.
Brookport, NY 14420

SW.CP.GL.frm.u150.0001
Strong West Lot to Lot Verification Worksheet

Strong West Lot to Lot Verification Worksheet

TECH: CLB

TEST: HC G

DATE: 11/30/2015

CURRENT REAGENT LOT# 604395

NEW REAGENT LOT # 616063

Current		
Sample Results	Instrument	Patients Order number
1	11.7	2 E3293189
2	19.39	2 E3291661
3	0.39	2 E3292816
4	1.35	2 E3292742
5	62.77	2 E3291440

New		
Sample Results	Instrument	
1	12.62	2
2	20.38	2
3	0.5	2
4	1.59	2
5	66.31	2

	% Difference	Delta Difference	Pass	Fail
1	-7.86324786	-0.92		
2	-5.1057246	-0.99		
3	-28.2051282	-0.11		
4	-17.7777778	-0.24		
5	-5.63963677	-3.54		

Current QC Range

Low Control	6.22 SD= .26
-------------	--------------

Current QC Range

High Control	45.48 SD= 1.26
--------------	----------------

Third level (if	SD=
-----------------	-----

New Lot QC Results

Low Control	6.48
-------------	------

High Control	48.4
--------------	------

Third Level	
-------------	--

Current Lot QC Results

Low Control	6.16
-------------	------

High Control	45.49
--------------	-------

Third Level	
-------------	--

%Difference = Current - new X 100

Current

Delta Difference = Current - new

Patient comparison acceptable with delta difference < 1 for results > 2.0 or +/- 0.3 for results < 2.0

AMM < 16% Difference

Reviewed by:

11/30/15