

TAYLOR HOSPITAL- CROZER

175 E Chester Pike
Ridley Park, PA 19078

Access2 - 2


Calibration Data Report
Access 2 Immunoassay System
S/N 510156, Version 3.6.0

SYS ID 9671706

Test Name:	PrCT	Curve:	Last Run; Active
Reagent Lot:	125321	Status:	Passed
Calibrated:	06/06/22 12:07	Reason:	
Expiration:	07/18/22 12:07	Cal Lot:	125314
Comment:			

	<u>Stated Conc. ng/mL</u>	<u>RLU</u>	<u>1 SD</u>	<u>%CV</u>	<u>Calculated Conc. ng/mL</u>	<u>Flags</u>
	S0 0.000	5946			No Value	
	S0 0.000	6165			0.000	
Mean		6055.365	154.496	2.551	0.000	
	S1 0.760	258304			0.854	
	S1 0.760	232574			0.777	
Mean		245439.000	18193.857	7.413	0.816	
	S2 4.850	1564340			4.329	
	S2 4.850	1579220			4.366	
Mean		1571780.000	10521.749	0.669	4.348	
	S3 10.110	4055670			10.399	
	S3 10.110	4024760			10.325	
Mean		4040215.000	21856.671	0.541	10.362	
	S4 26.010	10225400			25.788	
	S4 26.010	10297400			25.978	
Mean		10261400.000	50911.688	0.496	25.883	
	S5 47.140	17337700			46.325	
	S5 47.140	18305100			49.464	
Mean		17821400.000	684055.100	3.838	47.895	
	S6 98.100	29812500			96.988	
	S6 98.100	30002900			97.988	
Mean		29907700.000	134633.131	0.450	97.488	

Curve(Linear)



Technologist _____