





Specimen ID: 227621005329 1:20 Panels: CD
 Specimen Type: Whole blood Flag Set: Adult Priority: Routine
 Physician: Diagnosis:

Panels	Date	Time	Tube Pos	Instrument	Opr ID	Exceptions
CD	(S) 12/08/2021	16:38:23	99999	DxH1	ssexton	

Test	Result	Flags	Units
WBC	22.0	N	10 ³ /uL
UWBC	22.0	N	10 ³ /uL
RBC	0.03	N	10 ⁶ /uL
HGB	0.0	-	g/dL
HCT	0.6	R	%
MCV	189.1	+	fL
MCH	12.4	R	pg
MCHC	6.6	R	g/dL
RDW	62.4	+	%
PLT	62	N	10 ³ /uL
MPV	8.5	N	fL
NE	37.7	N	%
LY	18.5	N	%
MO	41.0	N	%
EO	2.0	N	%
BA	0.8	N	%
NE#	8.3	N	10 ³ /uL
LY#	4.1	N	10 ³ /uL
MO#	9.0	N	10 ³ /uL
EO#	0.4	N	10 ³ /uL
BA#	0.2	N	10 ³ /uL
NRBC	0.2	N	/100WBC
NRBC#	0.04	N	10 ³ /uL



*Jennifer Howard
12/08/21*

Actions: [1]:Check Sample Integrity, or Possible Aspiration Error.

Comment:

INF/DAT Filename: 2021-12-08T16-38-23_227621005329 1-20_99999_34307

Final Report
 Rev: 1.3.1.82 10.6402.6962 [3333]
 SYSTEM





Printed 12/08/2021 16:38:39

AD: R
 Page 1 of 1

Specimen ID: 227621005329 1:10 Panels: CD
 Specimen Type: Whole blood Flag Set: Adult Priority: Routine
 Physician: Diagnosis:

Panels	Date	Time	Tube Pos	Instrument	Opr ID	Exceptions
CD	(S) 12/08/2021	16:36:42	99999	DxH1	ssexton	

Test	Result	Flags	Units
WBC	44.7	N	10 ³ /uL
UWBC	44.7	N	10 ³ /uL
RBC	0.06	R	10 ⁶ /uL
HGB	0.1	-	g/dL
HCT		%
MCV	+++++		fL
MCH	10.2	R	pg
MCHC		g/dL
RDW		%
PLT	123	N	10 ³ /uL
MPV	8.8	N	fL
NE	29.9	N	%
LY	43.0	N	%
MO	26.7	N	%
EO	0.0	N	%
BA	0.4	N	%
NE#	13.4	N	10 ³ /uL
LY#	19.2	N	10 ³ /uL
MO#	11.9	N	10 ³ /uL
EO#	0.0	N	10 ³ /uL
BA#	0.2	N	10 ³ /uL
NRBC	0.1	N	/100WBC
NRBC#	0.04	N	10 ³ /uL



*Jennifer Howard
12/08/21*

Actions: [1]:Check Sample Integrity, or Possible Aspiration Error.

Comment:

INF/DAT Filename: 2021-12-08T16-36-42_227621005329 1-10_99999_34306

Final Report
 Rev: 1.3.1.82 10.6402.6962 [3333]
 SYSTEM

Printed 12/08/2021 16:37:02

AD: R
 Page 1 of 1