

Test Results



Redlands Community Hospital - 350 Terracina Blvd. Redlands, CA 92373
 Instrument ID: M20377 Instrument name: Immucor Echo
 Report generated: Sunday, May 5, 2019 09:41 AM

Test name	Group Screen	Software version	2.1.0.41
Operator	blood	Assay version	2.0U
Batch number	1350	Instrument ID	M20377
Profile name	Group Screen	Instrument name	Immucor Echo

Results - Summary

Sample ID	Group	Screen
	O Neg	Positive

Results - Group

Sample ID	Interp.	Flags	Mono Ctrl	Anti-A	Anti-B	Anti-D1	Anti-D2	A1 Cells	B Cells
	O Neg		0	0	0	0	0	4+	2+

Results - Screen

Sample ID	Interp.	Flags	Screen 1	Screen 2	Screen 3	Pos Ctrl
	Positive		3+	3+	0	4+

Supplies

Item name	ID	Lot:#	Serial#	QC	Expiration	Flags
Mono Control	492	341	1937639	05/01/2019	12/31/2019	
Anti-A	101	915	1480596	05/01/2019	08/14/2020	
Anti-B	203	771	0364400	05/01/2019	08/16/2020	
Anti-D Series 4	504	828	0963662	05/01/2019	10/18/2020	
Anti-D Series 5	505	821	0482528	05/01/2019	12/31/2019	
A1 Reference Cells	111	241	0171365	05/01/2019	05/24/2019	
B Reference Cells	113	241	0164581	05/01/2019	05/24/2019	
Capture LISS	211	181	0229810	05/01/2019	10/12/2019	
Capture R Indicators	221	305	0210878	05/01/2019	05/09/2019	
CMT strip	018	050	-	05/01/2019	11/08/2019	
Capture-R RS (3) strip	019	055	-	05/01/2019	06/25/2019	

Sample status

Sample ID	Started	Finished	Edited	Approved	Exported
249953	05/01/2019 08:57 PM blood	05/01/2019 09:22 PM blood	-	-	05/01/2019 09:22 PM blood

Legend

C	Control well failure	W	Weak D reflex
D	Manually edited crossmatch donor	X	Empty well
H	Hemolyzed or sample interference	?	Equivocal
M	Monolayer verification error	*	Manually edited
Q	QC failure	>	Volume too high
R	Reagent verification error	<	Volume too low
U	Undetermined failure		

Approved by: _____ Date: _____